

**Protocol for Analytical Methods
Used in the Assessment of
Properties under Part XV.1 of the
*Environmental Protection Act***

**Laboratory Services Branch
Ministry of the Environment**

March 9, 2004, amended as of July 1, 2011

«Cette publication hautement spécialisée n'est disponible qu'en anglais en vertu du règlement 441/97, qui en exempte l'application de la Loi sur les services en français. Pour obtenir de l'aide en français, veuillez communiquer avec le ministère de l'Environnement au 416 235-6348.»

TABLE OF CONTENTS

TABLE OF CONTENTS	1
ACRONYMS	3
GLOSSARY	5
SECTION 1: INTRODUCTION	8
1.1 Organic Parameter Groups	9
1.1.1 Acid/Base/Neutral Compounds (ABNs)*	9
1.1.2 Chlorophenols (CPs)*	9
1.1.3 1,4-Dioxane*	9
1.1.4 Dibenzo-p-Dioxins/Dibenzofurans (Dioxins/Furans, PCDDs/PCDFs)	9
1.1.5 Organochlorine Pesticides (OCs)	10
1.1.6 Petroleum Hydrocarbons (PHCs)	10
1.1.7 Polychlorinated Biphenyls (PCBs)	11
1.1.8 Polycyclic Aromatic Hydrocarbons (PAHs)	11
1.1.9 Trihalomethanes (THMs)*	11
1.1.10 Volatile Organic Compounds I (VOCs)	11
1.1.11 Volatile Organic Compounds II: Benzene, Toluene, Ethylbenzene, Xylene (BTEX)*	12
1.2 Inorganic Parameter Groups	12
1.2.1 Calcium and Magnesium (Ca, Mg)	12
1.2.2 Metals	12
1.2.3 Metals, Hydride-Forming (As, Se and Sb)*	13
1.2.4 Sodium (Na)	13
1.3 Other Regulated Parameters (ORPs)*:	13
SECTION 2: SAMPLE HANDLING AND STORAGE REQUIREMENTS	14
TABLE A: SOIL AND SEDIMENT Sample Handling and Storage Requirements	16
TABLE B: GROUND WATER Sample Handling and Storage Requirement	17
2.1 Subsampling:	20
2.1.1 Procedure: Soil and Sediment – Inorganic/Other Regulated Parameters	20
2.1.2 Procedure: Soil and Sediment – Organic Parameters	21
2.1.3 Procedure: Ground Water Samples – Inorganic/Other Regulated Parameters	21
2.1.4 Ground Water Samples – Organic Parameters	22
SECTION 3: ANALYTICAL METHODS	24
3.1 Analytical Method Summaries	24
3.1.1 Organic Parameters	25
3.1.1.1 Acid/Base/Neutral Extractable Organic Compounds (ABNs)	25
3.1.1.2 Chlorophenols (CPs)*	26
3.1.1.3 1,4-Dioxane*	27
3.1.1.4 Dibenzo-p-Dioxins/Dibenzofurans (Dioxins/Furans, PCDDs/PCDFs)	27
3.1.1.5 Organochlorine Pesticides (OCs)	30
3.1.1.6 Petroleum Hydrocarbons (PHCs)	31
3.1.1.7 Polychlorinated Biphenyls (PCBs)	34
3.1.1.8 Polycyclic Aromatic Hydrocarbons (PAHs) (may be analyzed with ABNs)	36
3.1.1.9 Trihalomethanes (THMs)*	37
3.1.1.10 Volatile Organic Compounds I (VOCs)	38
3.1.1.11 Volatile Organic Compounds II: Benzene, Ethylbenzene, Toluene, Xylenes (BTEX)*	40
3.1.2 Inorganic Chemical/Physical and Other Regulated Parameters	40
3.1.2.1 Boron (B-HWS) (Hot Water Soluble)	40
3.1.2.2 Calcium and Magnesium	41

3.1.2.3 Chloride (Cl ⁻) (water extractable)	41
3.1.2.4 Cyanide (CN ⁻)	42
3.1.2.5 Electrical Conductivity	43
3.1.2.6 Hexavalent Chromium (Chromium VI, Cr (VI), Cr ⁺⁶)	44
3.1.2.7 Mercury (Hg)	45
3.1.2.8 Methyl Mercury (Monomethyl Mercury, CH ₃ Hg ⁺ , MeHg ⁺)	45
3.1.2.9 Metals	46
3.1.2.10 Metals, Hydride-Forming (As, Se and Sb)*	47
3.1.2.11 pH by Potentiometry	48
3.1.2.12 Sodium (Na)	49
3.1.2.13 Sodium Adsorption Ratio (SAR)	50
3.1.2.14 Fraction of Organic Carbon (FOC)	50
SECTION 4: REPORTING	52
4.1 Required Reporting Limits (RLs)	52
TABLE 4.1.1 Required Reporting Limits	52
TABLE 4.1.2 PREVIOUS VALUES: 2004 Analytical Protocol APPENDIX B: Soil, Sediment and Water Standards and RLs	55
4.2: Reporting Requirements	59
4.3: Sample Dilution	60
4.3.1 Elevated non-target analyte or matrix interferences resulting in RLs above the standard	61
SECTION 5: REQUIRED QUALITY ASSURANCE/QUALITY CONTROL (QA/QC)	62
5.1 Accreditation:	62
5.2 Method Validation:	62
5.2.1 Demonstration of Acceptable Precision, Accuracy, Selectivity and Specificity:	62
5.3 Method Detection Limits:	63
5.3.1 Determination of MDL for Summed Parameters	65
5.3.2 Determination of MDL for Sodium Adsorption Ratio (SAR)	65
5.3.3 Calculation of Toxic Equivalence MDL	66
5.4 Measurement Uncertainty	67
5.5 Quality Control Samples	68
5.5.1 Acceptance Limits and Qualifiers	69
TABLE 5-1: Performance Criteria – Acid/Base Neutral Extractable Organic Compounds (ABNs), Chlorophenols (CPs), Polycyclic Aromatic Hydrocarbons (PAHs)	71
TABLE 5-2: Performance Criteria – 1,4-Dioxane	72
TABLE 5-3: Performance Criteria – Dioxins/Furans	73
TABLE 5-4: Performance Criteria – Organochlorine (OC) Pesticides	74
TABLE 5-5: Performance Criteria – Polychlorinated Biphenyls (PCBs)	75
TABLE 5-6: Performance Criteria – Petroleum Hydrocarbons (PHCs)	76
TABLE 5-7: Performance Criteria – Volatile Organic Compounds (VOCs)	77
TABLE 5-8: Performance Criteria – Cyanide (CN ⁻)	78
TABLE 5-9: Performance Criteria – Electrical Conductivity (EC)	79
TABLE 5-10: Performance Criteria – Fraction Organic Carbon (FOC), Chloride	80
TABLE 5-11: Performance Criteria – Hexavalent Chromium, Cr(VI)	81
TABLE 5-12: Performance Criteria – Mercury	82
TABLE 5-13: Performance Criteria – Methyl Mercury	83
TABLE 5-14: Performance Criteria – Boron, Hot Water Soluble (HWS); Calcium, Magnesium; Sodium; Metals (Including Hydride-Forming Metals)	84
TABLE 5-15 Performance Criteria – pH in Soil	85
SECTION 6: REFERENCES	86

ACRONYMS

AAS	atomic absorption spectrophotometry
ABN	acid base neutral extractable
ASTM	ASTM International (formerly the American Society for Testing and Materials)
B[a]P	Benzo[a]pyrene
BTEX	benzene/toluene/ethylbenzene/xylene
CALA	Canadian Association for Laboratory Accreditation (formerly the Canadian Association for Environmental Analytical Laboratories, CAEAL)
CAS	Chemical Abstract Service of the American Chemical Society
CCME	Canadian Council of Ministers of the Environment
CCV	continuing calibration verification
CP	chlorophenols
CRM	Certified Reference Material
CofA	Certificate of Analysis
CVAAS	cold vapour atomic absorption spectrophotometry
CVAFS	cold vapour atomic fluorescence spectrophotometry
DF	dilution factor
DLPCB	dioxin-like polychlorinated biphenyl
DNP	2,4-dinitrophenol
ECD	electron capture detector
EPA	<i>Environmental Protection Act</i> , R.S.O. 1990, c. E.19
FID	flame ionization detector
FOC	fraction organic carbon
GC	gas chromatography (or GLC, gas liquid chromatography)
GCxGC (or 2DGC)	two-dimensional gas chromatography
GC-ECD	gas chromatography-electron capture detector
GC-FID	gas chromatography-flame ionization detector
GC-HRMS	gas chromatography-high resolution mass spectrometry
GC-MS	gas chromatography-mass spectrometry
GC-MS/MS	gas chromatography-tandem mass spectrometry
GFAAS	graphite furnace atomic absorption spectrophotometry
HDPE	high density polyethylene
HGAAS	hydride generation atomic absorption spectrophotometry
HPLC-UV	high performance liquid chromatography with ultraviolet detector
HPLC-UV/FLU	high performance liquid chromatography with ultraviolet and fluorescence detectors
HRGC-HRMS	high resolution gas chromatography-high resolution mass spectrometry
HWSB	hot water soluble boron
ICP	inductively coupled plasma spectroscopy
ICP-OES	inductively coupled plasma-optical atomic emission spectroscopy
ICP-MS	inductively coupled plasma-mass spectrometry
IUPAC	International Union of Pure and Applied Chemistry
LSB, LaSB	Laboratory Services Branch
MDL	method detection limit
MOE	Ontario Ministry of the Environment
MS	mass spectrometry

OC	organochlorine pesticide
O. Reg. 153/04	Ontario Regulation 153/04 Records of Site Condition: Part XV.1 of the <i>Environmental Protection Act</i>
ORP	other regulated parameters
PAH	polycyclic aromatic hydrocarbon
PCB	polychlorinated biphenyl
PCDD	polychlorinated dibenzo-p-dioxin
PCDF	polychlorinated dibenzofuran
PET	polyethylene terephthalate
PHC	petroleum hydrocarbon
PT	performance testing (refers to performance testing sample)
QA	quality assurance
QC	quality control
QMS	quality management section
QP	qualified person
RL	reporting limit
RPD	relative percent difference
RSC	Record of Site Condition
SAR	sodium adsorption ratio
SIM	selected ion monitoring
SCC	Standards Council of Canada
SM	Standard Methods (American Public Health Association/American Water Works Association/Water Environmental Federation)
SVOC	semivolatile organic compound
TEF	toxic equivalency factor
TEQ	toxic equivalent
USEPA	United States Environmental Protection Agency
USGS-NWQL	United States Geological Survey-National Water Quality Laboratory
VOC	volatile organic compound

GLOSSARY

Accreditation: Formal recognition that a testing laboratory is competent to carry out specific tests or specific types of test.

Analyte: A substance or chemical constituent that is determined in an analytical procedure, such as a titration.

Analytical Run: A group of samples processed together through each step of an analytical procedure.

Analytical Standards: A series of chemical standards of the target analytes, used to set the relationship between instrument response and concentration or qualitative verification of instrument output.

Blank: Pure Water or other type of blank (i.e., acid or solvent) used to monitor for contaminated reagents, glassware and method processes.

Composite Sample: A sample that is made up of a number of laboratory grab samples from a single sample container that have been thoroughly mixed together.

Contaminant: Any solid, liquid, gas, odour, heat, sound, vibration, radiation or combination of any of them resulting directly or indirectly from human activities that may cause an adverse effect.

Certified Reference Material (CRM): A reference material, accompanied by documentation issued by an authoritative body and providing one or more specified property values with associated uncertainties and traceabilities using valid procedures (ISO/IEC GUIDE 99.2007).

Duplicate Sample: One of two samples taken from the same population and carried through all steps of sampling and analytical procedures in an identical manner.

Extractable Organic Compound: An organic compound which has a boiling point higher than water and which may vaporize when exposed to temperatures above room temperature: also known as a semivolatile organic compound (SVOC).

Field Blank: Blanks are defined as matrices that have negligible or unmeasurable amounts of the substance of interest. They are prepared by transferring the analyte free media from one vessel to another or by exposing the media to the sampling environment at the sampling site. Also known as a travel blank.

Field Filter: Where required, ground water samples must be filtered using a with a 0.45µm membrane filter (within 24 hours of sampling) and immediately preserved (if preservation is required).

Field Preserve: Where required, samples must be preserved with the specified preservative for that parameter group (within 24 hours of sampling) immediately following filtration (if filtration is required).

Hermetic Sampler: A commercially available, USEPA accepted device for sampling soil for VOC analysis. The device is inserted into the soil where it collects and seals a soil core (with no headspace). The device is transported to the laboratory where the entire sample is extracted and analyzed.

Holding Time: Elapsed time between sample collection and commencement of sample preparation or analysis, as appropriate.

Internal Standard: A standard that has chemical characteristics similar to those of the analyte(s) and provides an analytical response that is distinct from the analyte and not subject to interference. Internal standards are usually added to the sample or sample extract just prior to sample analysis in order to correct for variations in sample matrix, injection volume, etc.

ISO/IEC 17025 Standard: The requirements of the International Organization for Standardization, as amended from time to time, for testing laboratories to demonstrate that they are technically competent, maintain a quality system appropriate to the scope of their activities, and are able to generate technically valid calibration or test results.

Laboratory Control Sample: A sample of known concentration used as a basis for comparison with test samples, and which undergoes sample processing identical to that carried out for test samples. This sample is also referred to as a blank spike.

Laboratory Duplicate Sample: One of two sample aliquots obtained from the same sample container and carried through the entire analytical process. Also referred to as a split sample.

Matrix: The environment from which a given sample is taken (analytical chemistry) usually air, soil/sediment, ground or surface water for the purposes of this document.

Method Detection Limit (MDL): The minimum concentration of an analyte that can be identified, measured and reported with 99% confidence that the analyte concentration is greater than zero; it is determined from data produced by analysing a sample in a given matrix containing the analyte (CAN-P-1585-November 2006).

Method Blank: A blank sample which undergoes sample processing identical to that carried out for the test samples. Method blank results are used to assess contamination from the laboratory environment and reagents.

Method of Standard Additions: The determination of analyte concentration by adding known analyte amounts (spikes) to sample aliquots. Determination is based on the slope and intercept of the standard additions curve (recovery). The analytical response must be linear. The technique is used to correct for matrix effects.

Parameter: A parameter to be tested and is synonymous with other terminology such as “contaminant”, “target analyte”, or “analyte” which may be used in the regulation or other documents related to **O. Reg. 153/04**.

Qualified Person(s) (QP): A person as defined by **Ontario Regulation 153/04 Records of Site Condition: Part XV.1 of the *Environmental Protection Act***, Part II Defined Persons, s. 5 (as amended from time to time)

Quality Assurance (QA): Quality assurance is a system of planned activities intended to provide adequate confidence that quality requirements are being met. Quality assurance is one element of the quality system.

Quality Control (QC): Quality control is a set of operational techniques and activities intended to ensure that quality requirements are actually being met within known probability limits. Quality control is one part of the quality system.

Quality Control Sample: A sample (e.g., test sample or laboratory control sample/standard) used either singly or in replicate, as appropriate, to monitor performance characteristics [ISO 3534-1, 2.30].

Quality System: A set of interrelated elements (e.g., policies and objectives) that direct and control the way a facility operates with regard to quality.

Replicate Analyses: Natural samples may be split in the laboratory and analyzed together in the same run. Replicates are taken through the entire method process. This data can be used to assess the within-run precision of the analysis or sample matrix homogeneity.

Replicate Sample: An additional or second aliquot (portion) of a randomly selected sample in the analytical run.

Representative Sample: A subsample of material that has been taken so that it has essentially the same composition and characteristics of the sample in the container.

Reporting Limit (RL): The concentration at which a single analysis using the methods and matrices listed in this document will consistently detect target analytes when present. The RL must be equal to or greater than the MDL.

Reference Material (RM): A material or substance one or more of whose property values are sufficiently homogeneous and well established to be used for calibration of an apparatus, the assessment of a measurement method, or for assigning values to the materials. (ISO Guide 43-1). The RM should be matrix matched to the samples and carried through the entire analytical process.

Relative Percent Difference (RPD): The absolute difference between two results expressed as a percentage of the average result:

$$RPD = \left| \frac{(x_1 - x_2)}{(x_1 + x_2)/2} \right| \times 100$$

Significant Figures: The number of figures required to express a numerical determination such that only the last figure is uncertain. The number of significant figures, usually one or two for environmental tests is dependent upon method precision at the measured value.

Site Condition Standards: For the purpose of this protocol the prescribed contaminants and the site condition standards for those contaminants are those set out in Tables 1 through 9 of the **Soil, Ground Water and Sediment Standards. O. Reg. 153/04, s. 34 (1)**.

Surrogate: Has chemical characteristics similar to that of the analyte and provides an analytical response which is distinct from the analyte. The surrogate(s) is normally added to the sample prior to sample preparation and used to assess the recovery of analyte(s) carried through the analytical process.

Spiked Samples: Analyte(s) of interest is spiked into the sample matrix in order to monitor recovery from the sample matrix using the method or parts of the method.

Uncertainty: A non-negative parameter associated with the result of a measurement that characterizes the dispersion of the values that could reasonably be attributed to the measurand. (International vocabulary of metrology – Basic and general concepts and associated terms; ISO/IEC Guide 99:2007 (VIM 2007)).

Volatile Organic Compound (VOC): Any organic compound having, at 20 °C, a vapour pressure of 0.01kPa or more or having a corresponding volatility under the particular conditions of use, which is released into the atmosphere.

SECTION 1: INTRODUCTION

This protocol document is incorporated in the **Ontario Regulation 153/04 Records of Site Condition – Part XV.1 of the *Environmental Protection Act* (O. Reg. 153/04)**. It provides specific requirements for laboratory sample submission, analysis and data reporting.

Samples must be submitted to a laboratory that is accredited by an internationally recognized accreditation body [e.g., Standards Council of Canada (SCC), or Canadian Association for Laboratory Accreditation (CALA)] in accordance with the International Standard ISO/IEC17025:2005 – General Requirements for the Competence of Testing and Calibration Laboratories. Accreditation ensures that laboratories maintain a comprehensive documented quality system consistent with good analytical practice. Accreditation establishes a consistent basis for acceptable quality among analytical laboratories and ensures they adopt a satisfactory quality system to carry out sample analysis.

This protocol sets out the sample handling and storage requirements, analytical methods and method specific quality control and assurance procedures for laboratories established by recognized organizations: United States Environmental Protection Agency (USEPA), Ontario Ministry of the Environment (MOE) Laboratory Services Branch (LaSB), ASTM International (formerly American Society for Testing and Materials), Standard Methods: American Public Health Association (APHA)/American Water Works Association (AWWA)/Water Environmental Federation (WEF), U.S. Geological Survey (USGS) of the U.S. Department of the Interior, and the National Water Quality Laboratory (USGS-NWQL), Massachusetts Department of Environmental Protection Bureau of Waste Site Cleanup, Environment Canada and the Canadian Council of Ministers of the Environment (CCME).

The information in this protocol is provided to ensure that appropriate samples are submitted to laboratories, the samples are analyzed with methods that are fit for purpose and that the results of laboratory analyses are reported with sufficient quality upon which to base decisions required for **O. Reg. 153/04**.

Unless the wording of this protocol explicitly indicates that a statement is describing options, the words of the protocol are requirements which must be followed, subject to **O. Reg. 153/04** and other applicable law. Wording in this protocol which is explicitly mandatory, as well as wording in it which is simply descriptive, both set out requirements which must be followed.

Sample processing and analysis depends largely on the chemical and physical properties of the parameter to be measured. Parameters with similar chemical and physical properties can be grouped and processed together. Section 1 contains the parameters that can be grouped and processed together.

IMPORTANT: A laboratory is required to analyze and report all parameters listed within a parameter group (Section 1.1 and 1.2) when a parameter or parameters from that group are requested. Laboratories are allowed to combine parameter groups (e.g., parameter group 1.1.1 with 1.1.2 or parameter group 1.1.9 with 1.1.10 and 1.1.11) when processing samples, if the method in use has been accredited for those parameters and those parameters are listed in the reference method cited by the laboratory. Other regulated parameters (ORPs) listed in Section 1.3 can be reported individually.

Chemical Abstracts Service Registry Numbers (CAS RNs) for individual chemical parameters (where applicable) are listed in Table 4.1.1. Required Reporting Limits (Section 4).

1.1 ORGANIC PARAMETER GROUPS

1.1.1 Acid/Base/Neutral Compounds (ABNs)*

Parameters

biphenyl, 1,1-	dichlorobenzidine, 3,3'-	dinitrotoluene, 2,4-(2,6-) [#]
bis(2-chloroethyl)ether	diethyl phthalate	phenol
bis(2-chloroisopropylether)	dimethyl phthalate	trichlorobenzene, 1,2,4-
bis(2-ethylhexyl)phthalate	dimethylphenol, 2,4-	
chloroaniline, p	dinitrophenol, 2,4-	

*selected ABN parameters contained within **O. Reg. 153/04**

[#]the sum of 2,4- and 2,6-dinitrotoluene is compared to the standard

1.1.2 Chlorophenols (CPs)*

Parameters

chlorophenol, 2-	trichlorophenol, 2,4,5-
dichlorophenol, 2,4-	pentachlorophenol
trichlorophenol, 2,4,6-	

*may also be determined with ABNs

1.1.3 1,4-Dioxane*

Parameters

dioxane, 1,4-

*may also be determined with ABNs or VOCs

1.1.4 Dibenzo-p-Dioxins/Dibenzofurans (Dioxins/Furans, PCDDs/PCDFs)

Parameters

Congener Groups	2,3,7,8-Substituted Isomers
total tetrachlorodibenzo-p-dioxins (T4CDDs)	2,3,7,8-T4CDD
total pentachlorodibenzo-p-dioxins (P5CDDs)	1,2,3,7,8-P5CDD
total hexachlorodibenzo-p-dioxins (H6CDDs)	1,2,3,4,7,8-H6CDD
	1,2,3,6,7,8-H6CDD
	1,2,3,7,8,9-H6CDD
total heptachlorodibenzo-p-dioxins (H7CDDs)	1,2,3,4,6,7,8-H7CDD
octachlorodibenzo-p-dioxin (O8CDD)	1,2,3,4,6,7,8,9-O8CDD

total tetrachlorodibenzofurans (T4CDFs)	2,3,7,8-T4CDF
total pentachlorodibenzofurans (P5CDFs)	1,2,3,7,8-P5CDF
total hexachlorodibenzofurans (H6CDFs)	2,3,4,7,8-P5CDF
	1,2,3,4,7,8-H6CDF
	1,2,3,6,7,8-H6CDF
	1,2,3,7,8,9-H6CDF
	2,3,4,6,7,8-H6CDF
total heptachlorodibenzofurans (H7CDFs)	1,2,3,4,6,7,8-H7CDF
	1,2,3,4,7,8,9-H7CDF
octachlorodibenzofuran (O8CDF)	1,2,3,4,6,7,8,9-O8CDF

1.1.5 Organochlorine Pesticides (OCs)

Parameters (Synonym)

aldrin	endosulfan II (thiodan sulphate II) ²
chlordane, <i>alpha</i> - (α -chlordane) ¹	endrin
chlordane, <i>gamma</i> - (γ -chlordane) ¹	heptachlor
DDD ³	heptachlor epoxide
DDE ³	hexachlorobenzene
DDT ³	hexachlorobutadiene
dieldrin	hexachloroethane
hexachlorocyclohexane, <i>gamma</i> - (γ -HCH, lindane, γ -BHC*)	methoxychlor (DMDT)
endosulfan I (thiodan sulphate I) ²	

*erroneously known as benzene hexachloride (BHC)

¹the sum of *alpha*- and *gamma*-chlordane is compared to the standard

² the sum of endosulfan I and II is compared to the standard

³DDT standard applies to the total DDT (i.e., sum of the DDT isomers), the DDE standard applies to total DDE (i.e., sum of the DDE isomers), and the DDD standard applies to the total DDD (i.e., sum of the DDD isomers.)

1.1.6 Petroleum Hydrocarbons (PHCs)

Parameters

petroleum hydrocarbons (PHCs) (C₆–C₁₀ Fraction)

F1 (C₆ to C₁₀)

petroleum hydrocarbons (PHCs) (C₁₀–C₅₀ Fraction)

F2 (C₁₀ to C₁₆), F3 (C₁₆ to C₃₄), F4[#] (C₃₄ to C₅₀), F4G[#] (gravimetric)

[#]the larger result obtained for F4 and F4G is compared to the standard

1.1.7 Polychlorinated Biphenyls (PCBs)

Parameters

Aroclor 1242
 Aroclor 1248
 Aroclor 1254
 Aroclor 1260
 polychlorinated biphenyls (PCBs), total

1.1.8 Polycyclic Aromatic Hydrocarbons (PAHs)

Parameters (Synonym)

acenaphthene	benzo[g,h,i]perylene	indeno[1,2,3-cd]pyrene
acenaphthylene	benzo[k]fluoranthene	methylnaphthalene, 2-(1-) [#]
anthracene	chrysene	naphthalene
benz[a]anthracene	dibenz[a,h]anthracene	phenanthrene
benzo[a]pyrene (B[a]P)	fluoranthene	pyrene
benzo[b]fluoranthene	fluorene	

[#]the sum of 1- and 2-methylnaphthalene is compared to the standard

1.1.9 Trihalomethanes (THMs)*

Parameters (Synonyms)

bromodichloromethane (dichlorobromomethane)
 bromoform (tribromomethane)
 dibromochloromethane (chlorodibromomethane)

*may also be determined with VOCs

Note that the above list of compounds are commonly detected as a result of chlorination of drinking water and, therefore, are included as a separate group from the volatile organic compounds (Section 1.1.10)

1.1.10 Volatile Organic Compounds I (VOCs)

Parameters (Synonyms)

acetone (propanone)	dichloropropene, <i>trans</i> -1,3-*
bromomethane [#] (methyl bromide)	(dichloropropylene)
carbon tetrachloride (tetrachloromethane)	ethylene dibromide (dibromoethane, 1,2-)
chlorobenzene	hexane, n-**
chloroform (trichloromethane)	methyl ethyl ketone (MEK)
dichlorobenzene, 1,2-	methyl isobutyl ketone (MIBK)
dichlorobenzene, 1,3-	methyl <i>tert</i> -butyl ether (MTBE)
	methylene chloride (dichloromethane)

dichlorobenzene, 1,4-	styrene
dichlorodifluoromethane	tetrachloroethylene (tetrachloroethene, perchloroethylene)
dichloroethane, 1,1-	tetrachloroethane, 1,1,1,2-
dichloroethane, 1,2-	tetrachloroethane, 1,1,2,2-
dichloroethylene, 1,1- (dichloroethene)	trichloroethane, 1,1,1-
dichloroethylene, <i>trans</i> -1,2- (dichloroethene)	trichloroethane, 1,1,2-
dichloroethylene, <i>cis</i> -1,2- (dichloroethene)	trichloroethylene (trichloroethene)
dichloropropane, 1,2-	trichlorofluoromethane
dichloropropene, <i>cis</i> -1,3-*	vinyl chloride (chloroethene)
(dichloropropylene)	

*the sum of *cis*- and *trans*-dichloropropene is compared to the standard

#methanol-preserved samples may elevate the detection limit for bromomethane; a separate bisulphate-preserved sample or hermetically sealed sample may be submitted at the time of sampling if bromomethane is a chemical of concern.

**may also be determined with BTEX

1.1.11 Volatile Organic Compounds II: Benzene, Toluene, Ethylbenzene, Xylene (BTEX)*

Parameters (Synonyms)

benzene
ethylbenzene
toluene (methylbenzene)
xylenes, total (o-xylene; m- & p-xylene)

*may also be determined with VOCs

Note that the above BTEX compounds (benzene, toluene, ethylbenzene, xylenes) are a subset of volatile organic compounds (VOCs), are often analyzed as a discrete analysis and, therefore, are included as a separate group from the VOCs (Section 1.1.10).

1.2 INORGANIC PARAMETER GROUPS

1.2.1 Calcium and Magnesium (Ca, Mg)

Parameters

calcium (Ca)
magnesium (Mg)

1.2.2 Metals

Parameters

barium (Ba)	molybdenum (Mo)
beryllium (Be)	nickel (Ni)

boron (B)	silver (Ag)
cadmium (Cd)	thallium (Tl)
chromium (Cr)	uranium (U)
cobalt (Co)	vanadium (V)
copper (Cu)	zinc (Zn)
lead (Pb)	

1.2.3 Metals, Hydride-Forming (As, Se and Sb)*

Parameters

antimony (Sb)
arsenic (As)
selenium (Se)

*may also be determined with metals by ICP-MS or ICP-OES

1.2.4 Sodium (Na)

Parameters

sodium (Na)

1.3 OTHER REGULATED PARAMETERS (ORPs)*:

boron, hot water soluble (HWS)
chloride
cyanide
electrical conductivity
fraction of organic carbon (FOC)
hexavalent chromium
nitrate/nitrite
nitrogen, total
mercury
methyl mercury
pH
sodium adsorption ratio (SAR)

*The ORPs listed above are single parameter tests

Fraction organic carbon (FOC) is calculated from total organic carbon.

SECTION 2: SAMPLE HANDLING AND STORAGE REQUIREMENTS

This section provides details on the procedures for sample handling and storage, including the type of container, sample volume, preservation and storage requirements, and maximum holding time for all regulated analytes.

It is especially important for samples requiring organic analysis, that samples be placed in the appropriate containers and the cooling begun as soon as possible after sampling. Sufficient ice or other coolant should be added to produce a temperature less than (<) 10 °C. Note that samples arriving at the laboratory on the day of sampling may not have had time to achieve a temperature of < 10 °C. This is acceptable as long as the cooling process has begun.

Ground Water Sample Containers:

Extractable organics testing ground water samples is “whole bottle” analysis where the entire sample is extracted and the bottle rinsed with solvent. This is necessary to prevent analyte losses due to adsorption on the container walls. Thus, additional containers are required for laboratory QC (duplicates and matrix spikes). Similarly, for volatile organics testing, additional vials are required for laboratory QC and possible repeats because once a vial has been sampled it is not suitable for further testing. Consult the laboratory for the correct number of sample vials.

Most inorganic tests have differing container and preservative requirements. Chloride and electrical conductivity can be determined from a single unpreserved sample.

Ground Water Samples Requiring PAH Analysis:

Polycyclic aromatic hydrocarbons adsorb strongly to particulate matter. Thus, analysis of a ground water sample containing particulate may be biased high relative to the PAH actually dissolved in the water. If the QP notices particulate in a sample they have the option of collecting an additional sample for laboratory filtration prior to analysis for benzo[a]pyrene only.

Ground Water Samples Requiring Hexavalent Chromium Analysis:

For dissolved hexavalent chromium in ground water, the samples are field filtered through a 0.45 µm membrane filter, and the pH is adjusted to 9.3 to 9.7, with the addition of a buffer solution within 24 hours of sampling. **IMPORTANT:** In order to achieve the 28-day holding time specified in Table B, the samples must be preserved with the ammonium sulfate buffer solution specified in EPA Method 218.6 (revision 3.3, 1994) or Standard Methods 3500-Cr Chromium (2009). Unpreserved samples and/or samples preserved only with sodium hydroxide must be analyzed within 24 hours from sampling.

Soil and Sediment Sample Containers:

For organic compound testing, each analysis requires about 10 g of sample, thus, multiple tests can be conducted on a full 125 or 250 mL soil container. A separate container, usually 40–60 mL, is required for volatile organics testing. A single container will normally suffice for inorganic tests.

Soil and sediments samples requiring analysis for VOCs, BTEX, PHC (F1), THMs and 1,4-dioxane are preserved in the field with methanol or collected using hermetically sealed sampling devices. For BTEX and PHC (F1), this is an accepted deviation from the CCME method. An

additional sample collected in a glass jar is required for moisture content determination. Each batch of methanol-preserved soil samples require an additional vial pre-charged with methanol for the field/travel blank.

Note: Tables A and B and the notations below them provide both the sampler and those receiving samples at a laboratory with the requirements for sample container, sample preservation, sample storage and sample holding times. The number of sampling containers and container sizes specified is a guide. Always consult the laboratory prior to sampling. The laboratory will provide sufficient appropriate containers for the required scope of testing. Collection of multiple sample containers is encouraged to avoid the need for resampling if the sample is consumed or compromised during shipping and/or analysis.

TABLE A: SOIL AND SEDIMENT Sample Handling and Storage Requirements

SOIL Inorganic Parameters	Container ¹	Field Preservation	Storage Temp. ²	Preserved Holding Time ³	Unpreserved Holding Time ³
Chloride, electrical conductivity	glass, HDPE or PET	none	5 ± 3 °C		30 days as received (without lab drying); indefinite when dried at the lab
Cyanide (CN ⁻)	glass wide-mouth jar, Teflon™ lined lid	protect from light	5 ± 3 °C		14 days
Fraction organic carbon (FOC)	glass jar, Teflon™ lined lid	none	5 ± 3 °C		28 days as received(without lab drying); indefinite storage time when dried
Hexavalent chromium	glass, HDPE	none	5 ± 3 °C		30 days as received
Metals (includes hydride-forming metals, SAR, HWS boron, calcium, magnesium, sodium)	glass, HDPE	none	5 ± 3 °C		180 days as received (without lab drying); indefinite when dried at the lab
Mercury, methyl mercury	glass, HDPE or PET	none	5 ± 3 °C		28 days
pH	glass, HDPE or PET	none	5 ± 3 °C		30 days as received
SOIL Organic Parameters	Container ^{1,5,6,7,20}	Field Preservation	Storage Temp. ²	Preserved Holding Time ³	Unpreserved Holding Time ³
BTEX ⁸ , PHCs (F1) ⁸ , THMs, VOCs ⁷	40–60 mL glass vial (charged with methanol preservative, pre-weighed) ⁶ AND glass jar (for moisture content) [hermetic samplers are an acceptable alternative ^{5, 18}]	methanol (aqueous NaHSO ₄ is an acceptable alternative for bromomethane) ^{6, 7, 18, 20}	5 ± 3 °C	14 days	hermetic samples: stabilize with methanol preservative within 48 hours of sampling ¹⁸
NB: SEE FOOTNOTE #20					
1,4-Dioxane ^{9, 15}	when processed as a VOC sample: same as per VOCs above; when processed as an extractable: same as per ABNs below; (consult laboratory) ^{9, 15, 18}		5 ± 3 °C	14 days	when processed as a VOC sample: same as per VOCs above; when processed as an extractable: same as per ABNs below; (consult laboratory) ¹⁸
PHCs (F2–F4)	glass wide-mouth jar, Teflon™ lined lid	none	5 ± 3 °C		14 days
ABNs, CPs, OCs, PAHs	glass wide-mouth jar, Teflon™ lined lid	none	5 ± 3 °C		60 days
Dioxins and furans, PCBs	glass wide-mouth jar, Teflon™ lined lid	none	5 ± 3 °C		indefinite storage time

HDPE = high density polyethylene; PET = polyethylene terephthalate; HWS = hot water soluble boron; THM = trihalomethanes; VOC = volatile organic compounds; BTEX = benzene, toluene, ethylbenzene, xylenes; PHCs = petroleum hydrocarbons; CPs = chlorophenols; PCBs = polychlorinated biphenyls; OCs = organochlorine pesticides

^{1–20} footnotes immediately follow Table B

TABLE B: GROUND WATER Sample Handling and Storage Requirement

GROUND WATER Inorganic Parameters	Container ¹⁰	Field Preservation	Storage Temperature ²	Preserved Holding Time ³	Unpreserved Holding Time ³
Chloride, electrical conductivity, pH	HDPE or glass	none	5 ± 3 °C	14 days	28 days
Cyanide (CN ⁻)	HDPE or glass	NaOH to a pH > 12	5 ± 3 °C	14 days	must be field preserved
Hexavalent chromium	HDPE or glass	field filter followed by buffer solution to a pH 9.3–9.7 ¹⁷	5 ± 3 °C	28 days ¹⁷	24 hours ¹⁷
Metals (includes hydride-forming metals, calcium, magnesium, sodium)	HDPE or Teflon™ ¹⁰	field filter followed by HNO ₃ to pH < 2 ¹¹	room temperature when preserved	60 days	must be field preserved
Mercury	glass or Teflon™ ¹⁰	field filter followed by HCl to pH < 2 ¹¹	room temperature when preserved	28 days	must be field preserved
Methyl mercury	glass or Teflon™	DO NOT FILTER HCl or H ₂ SO ₄ to pH < 2 ¹²	5 ± 3 °C	28 days	DO NOT FILTER must be field preserved ¹²
GROUND WATER Organic Parameters^{10, 13, 14}	Container^{10, 13, 14}	Field Preservation	Storage Temperature²	Preserved Holding Time³	Unpreserved Holding Time³
BTEX, PHCs (F1), THMs, VOCs;	40–60 mL glass vials (minimum of 2) ¹⁴ (no headspace)	NaHSO ₄ or HCl to a pH < 2 ¹⁶	5 ± 3 °C	14 days	7 days
1,4-Dioxane ^{9, 15}	when processed as a VOC sample: same as per VOCs above; when processed as an extractable: same as per ABNs below; (consult laboratory) ^{9, 15}		5 ± 3 °C	14 days	14 days
PHCs (F2–F4)	1L amber glass bottle, Teflon™ lined lid	NaHSO ₄ or HCl to a pH < 2 ¹⁶	5 ± 3 °C	40 days	7 days
ABNs, CP, OCs, PAHs ¹⁹ , PCBs	1L amber glass bottle, Teflon™ lined lid	none	5 ± 3 °C		14 days
Dioxins and furans	1L amber glass bottle, Teflon™ lined lid	None	5 ± 3 °C		indefinite storage time

HIDPE = high density polyethylene; THM = trihalomethanes; VOC = volatile organic compounds; BTEX = benzene, toluene, ethylbenzene, xylenes; PHCs = petroleum hydrocarbons; CPs = chlorophenols; PCBs = polychlorinated biphenyls; OCs = organochlorine pesticides

¹ One soil container is generally sufficient for inorganic analysis and another for extractable organics. A separate container is required for BTEX, THM, VOC and PHC (F1) moisture analysis.

² Storage temperature refers to storage at the laboratory. Samples should be cooled and transported as soon as possible after collection.

³ Holding time refers to the time delay between time of sample collection and time stabilization/analysis is initiated. For samples stabilized with methanol, the hold time for the recovered methanol extract is up to 40 days.

- 4 PET can not be used for samples requiring antimony analysis.
- 5 As an alternative, the USEPA has investigated hermetic sample devices that take and seal a single core sample. The sample is submitted as is to the laboratory where it is extruded into an extracting solvent. Samples must be received at the laboratory within 48 hours of sampling. (Note that replicate samples are necessary for bisulphate and methanol extraction for all samples plus laboratory duplicates and spikes.) Consult the laboratory for the number of samples required.
- 6 The USEPA has approved field preservation. Pre-weighed vials containing known weights of methanol preservative (or aqueous sodium bisulphate if used for bromomethane) are sent to the field. Sample cores (approximately 5 g) are extruded directly into the vial. The vials are sealed, and submitted directly to the laboratory. In practice, this technique requires great care to prevent losses of methanol due to leaking vials or through splashing. Consult the laboratory for the number of containers required.
- 7 Methanol-preserved samples may elevate the detection limit for bromomethane (VOC); a separate bisulphate-preserved sample or hermetically sealed sample may be submitted at the time of sampling if bromomethane is a chemical of concern – contact the laboratory to determine if a separate sample should be collected.
- 8 For BTEX and PHC (F1) pre-charging the soil sampling container with methanol preservative is an accepted deviation from the CCME method.
- 9 1,4-Dioxane may be analyzed with the ABNs or VOCs; sample container requirements used for ABNs or VOCs are both acceptable. If 1,4-dioxane is to be analyzed with ABNs, follow the ABN sample container requirements; similarly if it is to be analyzed with VOCs, follow VOC sample container requirements. Consult the laboratory for the container type and the total number required (see also footnote #15).
- 10 Samples containing visual sediment at the time of analysis should be documented and noted on the Certificate of Analysis or written report as results may be biased high due to the inclusion of sediment in the extraction.
- 11 Field filter with 0.45µm immediately prior to adding preservative or filling pre-charged container.
- 12 Sample directly into a HCl or H₂SO₄ preserved container, or add acid to an unfiltered sample immediately after sample collection in the field.
- 13 Aqueous organic samples should be protected from light. If amber bottles are not available, glass should be wrapped in foil.
- 14 Separate containers are required for each organic water analysis. Consult the laboratory for required volumes. Chloride and electrical conductivity can be taken from the same container.
- 15 For 1,4-dioxane in soil and sediment, no preservative is required if processed as an ABN, however. Methanol is an acceptable alternative if processed as a VOC. For 1,4-dioxane in groundwater, no preservative is required, however, NaHSO₄ or HCl are acceptable alternatives.
- 16 Preserved to reduce biodegradation, however effervescence/degassing may occur in some ground water samples. In this case, rinse preservative out three times with sample and submit to the laboratory as unpreserved.
- 17 To achieve the 28-day holding time, use the ammonium sulfate buffer solution [i.e., (NH₄)₂SO₄/NH₄OH] or (NH₄)₂SO₄/NH₄OH/NaOH + NaOH] as specified in EPA Method 218.6 (revision 3.3, 1994) or Standard Methods 3500-Cr Chromium (2009). Using only NaOH without the ammonium sulfate buffer to adjust the pH would require analysis within 24 hours of sampling.
- 18 Alternatively, to achieve a longer hold time, hermetic samples may be frozen within 48 hours of sampling as per ASTM method D6418 – 09, however, storage stability must be validated by the laboratory with no more than 10% losses.
- 19 For benzo(a)pyrene in ground water samples filtration prior to analysis on a duplicate sample is permitted.
- 20 For VOC, BTEX, F1 PHCs, 1,4 dioxane soil samples collected before July 1, 2011, the following sampling and handling requirements are also permitted.

SOIL Organic Parameters	Container	Preservative	Storage Temperature	Preserved Holding Time	Unpreserved Holding Time
VOC, BTEX, F1 PHCs, 1,4-dioxane*	glass jar, Teflon lined lid, no headspace, separate container required Hermetic samplers are an acceptable alternative	none field preservation with aqueous sodium bisulphate and methanol is an acceptable alternative	5 ± 3C	See notations 1-3 below	Stabilize by extraction or freezing within 48 hrs of receipt at the laboratory (7days from sampling). Frozen or field preserved samples must be extracted within 14 days of sampling.

*Special care must be used when sampling for VOC, BTEX and F1 in soil and sediment. Studies have shown that substantial losses can occur through volatilization and bacterial degradation. There are several allowable options for field collection of samples. Each is discussed below. Consult SW846, Method 5035A for additional detail. The laboratory is required to stabilize the sample on the day of receipt, either by extraction or freezing.

1. Collection in soil containers: To minimize volatilization losses, minimize sample handling and mixing during the process of filling the sample container. The bottle should be filled with headspace and voids minimized. Care is required to ensure that no soil remains on the threads of the jar, preventing a tight seal and allowing volatilization losses. To minimize losses through bacterial degradation, commence cooling of the samples immediately and transport the samples to the lab as soon as possible, ideally on the day of sampling. Samples must be received at the laboratory within 48 hours of sampling. Freezing can be used to extend the hold time to 14 days, however the practice is difficult to implement in the field and can cause sample breakage.
2. As an alternative, the USEPA has investigated hermetic sample devices that take and seal a single core sample. The sampler is submitted as is to the laboratory where it is extruded into the extracting solvent. Samples must be received at the laboratory within 48 hours of sampling. This technique minimizes volatilization losses and is worth consideration for critical sites. (Note that replicate samplers are necessary for bisulphate and methanol extraction for all samples plus lab duplicates and spikes). Consult the laboratory for the number of samplers required.
3. The USEPA has also approved field preservation. Pre-weighed vials containing known weights of methanol and aqueous sodium bisulphate preservative are sent to the field. Sample cores (≈ 5 g) are extruded directly into the vial. The vials are sealed, and submitted directly to the laboratory. In practice, this technique requires great care to implement successfully. Losses due to leaking vials, through splashing and effervescence (aqueous bisulphate) can easily occur and make the sample unusable. Consult the laboratory for the number of containers required.

2.1 SUBSAMPLING:

The procedures described cover common situations when subsampling solid and liquid samples in the laboratory. When such situations arise these procedures shall be followed. All actions taken to obtain representative samples other than those described below must be included in the Certificate of Analysis or written report so that the Qualified Person (QP) will be able to properly assess the data and be able to determine if the data are of sufficient quality upon which to base decisions required for **O. Reg. 153/04**.

2.1.1 Procedure: Soil and Sediment – Inorganic/Other Regulated Parameters

1. Prior to homogenization or drying, samples are to be inspected in the laboratory for multiphase conditions (free water, petroleum product, etc.) or other anomalies. Small amounts of free water or petroleum product may be mixed with the sample but large amounts of free water or petroleum product should be separated. The QP should be contacted and agreement reached on how to proceed. All such anomalies and the actions taken must be noted in the Certificate of Analysis or analytical report.

Drying may change the pH of the soil; therefore, pH is conducted on the sample as received. Also, because of potential volatilization losses for cyanide and possible redox reactions for hexavalent chromium, samples for these tests are also taken from the sample as received.

The sample is mixed as well as possible and several aliquots taken to obtain the desired weight. Hard clay samples that cannot be mixed are “cored”, using a spatula in different spots or sections of the jar. Stones, twigs and other foreign materials are excluded. To ensure a representative subsample is obtained, a minimum 10 g aliquot is taken.

2. For other inorganic soil and sediment tests, samples are air or oven dried at a temperature of $\leq 60^{\circ}\text{C}$ to prevent the potential loss of volatile analytes, for a minimum of 48 hours or until no visible moisture remains. To determine moisture content a separate aliquot is taken and dried at 105°C overnight or until a constant weight is achieved.

Stones, twigs and other foreign materials are excluded from the subsamples.

Physical reduction of large clay aggregates is required.

Samples are then passed through a 2 mm sieve. Any portion that does not pass through this sieve is discarded. Minimum 5 g aliquots of the 2 mm portion of the sample are then used in the analysis of chloride, electrical conductivity, sodium absorption ratio and hot water soluble boron.

3. A subsample of the 2 mm portion is then taken and ground to pass through a $355\ \mu\text{m}$ sieve in its entirety. This portion is then used in the analysis of all metal parameters including hydride-forming metals mercury and fraction organic carbon.

Fraction Organic Carbon (FOC):

A minimum of three 5 g subsamples are required for each soil sample requiring FOC analysis; samples are analyzed and reported in triplicate.

2.1.2 Procedure: Soil and Sediment – Organic Parameters

1. Prior to subsampling, samples are inspected for multiphase conditions (free water, petroleum product, etc.) or other anomalies. Small amounts of free water or petroleum product may be mixed with the sample but large amounts of free water or petroleum product should be separated. The QP should be contacted and asked how to proceed. All such anomalies and the actions taken must be noted in the Certificate of Analysis.
2. Samples requiring organic analysis for ABN, dioxins, CP, OC pesticides, PAH and PCB are air dried for a minimum of 48 hours, or until no visible moisture remains, ground, and homogenized. Alternately to drying, samples can be mixed with equal amounts of anhydrous sodium sulphate, or until the sample resembles a free flowing powder. Stones, twigs and other foreign materials are excluded from the subsamples.
3. Field preserved samples requiring analysis for volatile analytes (VOC; BTEX; PHC (F1), THM) are processed as received. Samples collected in hermetic sampling devices are extruded directly into the extraction solvent.
4. Samples requiring analysis for PHC (F2, F3, F4 and F4G) are not field preserved. The use of sodium sulphate as a drying agent could lead to an exothermic reaction and thus should not be used. A minimum of 5 grams dry weight of the soil is taken for analysis. The extraction fluid is added immediately after weighing to minimize volatilization losses.
5. For all other organic analyses, the sample is mixed as well as possible and several aliquots taken to obtain the desired weight. Hard clay samples that cannot be mixed are “cored”, using a spatula in different spots or sections of the jar. Stones, twigs and other foreign materials are excluded. The extraction fluid should be added as soon as possible after weighing to minimize sample degradation.

2.1.3 Procedure: Ground Water Samples – Inorganic/Other Regulated Parameters

Prior to subsampling, samples are inspected for particulate and the approximate amount of visible particulate (v/v) noted. If particulate is > 5% v/v the QP is contacted. It may be necessary to separate the solids and treat as separate samples. If multiphase samples are encountered, (usually petroleum product on the surface), the non-aqueous phase is excluded from any subsamples. This is also noted and reported.

Electrical Conductivity and pH:

Do not shake, dilute or alter the samples in any way as this can alter the result. Pour the sample into the sample cup or measurement vessel.

Chloride, Cyanide:

Shake and pour the sample. Alternately an aliquot may be syringe filtered or decanted to prevent instrument problems.

Hexavalent Chromium:

Samples requiring analysis for hexavalent chromium in ground water are field filtered through a 0.45 µm filter immediately followed by field preservation as described under “Ground Water Samples

Requiring Hexavalent Chromium Analysis” at the beginning of Section 2. Unpreserved samples must be preserved by the laboratory within 24 hours of sampling. Unfiltered, preserved samples are not suitable for laboratory filtration. The filter media must be proven clean (i.e., analyte of interest is below the method detection limit).

Metals, Including Hydride-Forming Metals:

Ground water sample analysis for metals must be carried out on the dissolved fraction.

Dissolved Metals, Including Mercury:

Samples requiring analysis for dissolved metals and mercury in ground water are field filtered through a 0.45 µm filter immediately followed by field preservation. Unfiltered, preserved samples are not suitable for laboratory filtration. The filter media must be proven clean.

2.1.4 Ground Water Samples – Organic Parameters

Volatile Organic Compounds

Volatile organic compound samples (e.g., VOC; BTEX; PHC (F1); THM; 1,4-dioxane) are treated differently than extractable organic samples. Samples should be received in replicate VOC vials.

1. When sampling, the vials or bottles should be filled slowly to the top rim of the container so that a dome or convex meniscus is present. A slight loss of sample may occur when the cap is applied. When capped, the cap or septum should be in contact with the sample so that no air is trapped in the sample container and when the vial or bottle is turned upside down, any air bubble present should not cover the bottom of the vial. The Teflon™ liner, not the silicone or rubber backing of the septum, must be in contact with the sample
2. Prior to analysis, samples are inspected for particulate and the approximate amount of visible particulate (v/v) noted. They are also examined for headspace and if on inversion, there is an air bubble present that covers the bottom of the vial, the sample is compromised and should not be analyzed. If the client requires analysis, the data must be reported qualified.
3. Modern purge & trap autosampler systems permit direct aliquoting and surrogate / internal standard addition without opening the vial.
4. For older purge & trap apparatuses and headspace systems, the vial is opened and the appropriate aliquot is removed, immediately placed in the analysis vessel and the vessel sealed. Once opened, the sample vial is compromised and not suitable for reanalysis.
5. If the sample contains a non-aqueous layer, it is generally unsuitable for analysis. If the client requires analysis an aliquot may be removed with a syringe from below the non-aqueous layer, analyzed and the data qualified.

Extractable Organic Compounds

Extractable organic analytes tend to be hydrophobic and will adsorb to both the sample bottle and any particulate in the sample. As such, the default method of analysis is “whole sample” analysis in which the entire contents of the sample are extracted, the sample bottle rinsed with solvent and the combined extract used for analysis. Inclusion of particulates will tend to produce a high bias.

1. Prior to extraction the sample is inspected for particulate and a non-aqueous phase. If there is no non-aqueous phase, the amount of particulate (if any) is noted and reported on the Certificate of Analysis or analytical report, the entire sample is extracted, the sample bottle rinsed with solvent and the combined extract used for analysis.
2. If a surface “sheen” is observed it is noted and reported but the sample is treated as in number 1 above.
3. If a substantial (separable) non-aqueous layer is observed, the QP is contacted for instructions as to how to proceed. If instructions are not received, the non-aqueous layer is separated from the aqueous layer, its volume estimated and it is retained for possible analysis. The aqueous layer is extracted as in number 1 above.

There is one exception to the above procedure for PAH analysis of ground water samples: At the time of collection, if the QP notices particulate in the sample, he has the option of collection of a second sample which is submitted to the laboratory for filtration prior to analysis. The sample must be clearly labelled “for filtration and benzo[a]pyrene analysis”.

This sample is filtered through an inert medium and collected in a second sample bottle. Both the filtered and unfiltered samples are extracted as described in number 1 above.

SECTION 3: ANALYTICAL METHODS

The analytical methods described in this section derive from the following sources:

- ASTM International (formerly American Society for Testing and Materials) (www.astm.org)
- Canadian Council of Ministers of the Environment (CCME) (www.ccme.ca)
- Environment Canada (www.ec.gc.ca)
- Massachusetts Department of Environmental Protection Bureau of Waste Site Cleanup (<http://www.mass.gov/dep/cleanup>)
- Ontario Ministry of the Environment (MOE) Laboratory Services Branch (LaSB); email requests for these methods can be sent to LaboratoryServicesBranch@ontario.ca
- Standard Methods for the Examination of Water and Wastewater: American Public Health Association (APHA)/American Water Works Association (AWWA)/Water Environmental Federation (WEF) (www.standardmethods.org)
- United States Environmental Protection Agency (USEPA) (www.epa.gov)
- United States Geological Survey (USGS) of the United States Department of the Interior, (www.usgs.gov) and the National Water Quality Laboratory (USGS-NWQL) (<http://nwql.usgs.gov>)

Laboratories are required to verify that all procedures in the analytical method are documented and based on the latest valid edition of the reference method. Modifications to the analytical method are limited to the instructions provided in the method principle outlined below. All modifications to the analytical method must be documented; the method must be validated and must contain a statement that the method is fit for the intended use with respect to the sensitivity, selectivity, analytical range, and the method precision and bias.

All the method validation, quality assurance and quality control requirements in Section 5 must be met.

3.1 ANALYTICAL METHOD SUMMARIES

The methods are organized according to the type of parameter for which each method may be used.

There are several cases below where it is stated test groups can be analyzed together. Combining test groups may compromise analytical conditions. Such combinations are permitted only when all of the required performance standards in Tables 5-1 through to 5-15 are met.

3.1.1 ORGANIC PARAMETERS

3.1.1.1 Acid/Base/Neutral Extractable Organic Compounds (ABNs)

Parameters

biphenyl, 1,1-	dichlorobenzidine, 3,3'-	dinitrotoluene, 2,4-(2,6-) [#]
bis(2-chloroethyl)ether	diethyl phthalate	phenol
bis(2-chloroisopropylether)	dimethyl phthalate	trichlorobenzene, 1,2,4-
bis(2-ethylhexyl)phthalate	dimethylphenol, 2,4-	
chloroaniline, p-	dinitrophenol, 2,4-	

*selected ABN parameters contained within **O. Reg. 153/04**

[#]the sum of 2,4- and 2,6-dinitrotoluene is compared to the standard

Table 3.1.1.1

METHOD REFERENCE SOURCE	SOIL & SEDIMENT	GROUND WATER
USEPA	<p>Sample Preparation SW-846, Method 3540C SW-846, Method 3541 SW-846 Method 3546 SW-846, Method 3550C SW-846, Method 3570</p> <p>Sample Cleanup SW 846 Method 3610B SW846 Method 3630C</p> <p>Analysis SW-846, Method 8270D EPA Method 1625C</p>	<p>Sample Preparation SW-846, Method 3510C SW-846, Method 3520C SW-846, Method 3535A</p> <p>Analysis SW-846, Method 8270D EPA Method 1625C</p>
Standard Methods		Method 6410B
MOE-LaSB		E3265

Method principle

Aqueous or soil samples, as received, are fortified with surrogates and extracted with a solvent or solvent mix (Table 3.1.1.1).

Soil and sediment samples may be dried by mixing with a desiccant prior to extraction. Ground water sample extraction must be carried out at pH < 2 (acid extractables) and > 11 (base neutral extractables). Extracts are dried, concentrated and exchanged into a solvent compatible with the cleanup (if necessary) or determinative technique being employed.

Internal standards are added after all preparation and cleanup steps are completed. Extracts can be kept for up to 40 days. Analysis is by gas chromatography-mass spectrometry (GC-MS) operated in either the full scan or selected ion monitoring (SIM) mode. The SIM mode provides lower detection limits while the full scan mode provides diagnostic capability and permits investigation of non-target analytes.

Derivatization when required involves a chemical reaction that converts the phenol and chlorophenol analytes of interest to their corresponding esters resulting in improved chromatography and detection limits. Cleanup techniques remove interferences that may impact quantitation and degrade column performance. In general, derivatization will not be required to achieve the required reporting limits (RLs) for phenols and chlorophenols. Cleanup may be required for difficult samples and laboratories may elect to perform cleanups routinely to extend column life.

Quantitation is by the internal standard method.

3.1.1.2 Chlorophenols (CPs)*

Parameters

chlorophenol, 2-
dichlorophenol, 2,4-
trichlorophenol, 2,4,6-

trichlorophenol, 2,4,5-
pentachlorophenol

*may also be determined with ABNs

Table 3.1.1.2

METHOD REFERENCE SOURCE	SOIL & SEDIMENT	GROUND WATER
USEPA	Sample Preparation SW-846, Method 3540C SW-846, Method 3541 SW-846, Method 3546 SW-846, Method 3550C SW-846, Method 3570 Analysis SW-846, Method 8270D	SW-846, Method 3510C SW-846, Method 3520C SW-846, Method 3535A Analysis SW-846, Method 8270D
Standard Methods		Method 6420C 6410B
MOE-LaSB		E3119 E3265

Method principle

Reference methods in Table 3.1.1.2 shall be followed with the following addition: ground water samples must be acidified to pH < 2 prior to liquid/liquid extraction in order to achieve adequate recoveries. Solid phase extraction procedures (Method 3535A) may not require acidification.

3.1.1.3 1,4-Dioxane***Parameters**

dioxane, 1,4

*may also be determined with ABNs or VOCs)

Table 3.1.1.3

METHOD REFERENCE SOURCE	SOIL & SEDIMENT	GROUND WATER
USEPA	<p>Sample Introduction for 8260 SW-846 Method 5021 SW-846 Method 5035</p> <p>Sample Preparation for 8270 SW-846, Method 3540C SW-846, Method 3541 SW-846 Method 3546 SW-846, Method 3550C</p> <p>Analysis EPA 1624C EPA 1625C</p>	<p>Sample Introduction for 8260 SW-846 Method 5000 SW-846 Method 5030B</p> <p>Sample Preparation for 8270 SW-846, Method 3510C SW-846, Method 3520C SW-846, Method 3535A</p> <p>Analysis EPA 1624C EPA 1625C</p>

Method principle

1,4-Dioxane is a water soluble organic compound which can be analyzed either as an extractable organic or volatile organic compound

Because 1,4-dioxane recovers (or purges) poorly, isotope dilution, where the native analyte is quantitated using the deuterated analog (EPA1624C, EPA 1625) is required. Other than quantitation by isotope dilution, all other facets of the preparation and analysis are similar to the analysis of VOC (Table 3.1.1.10) or ABN (Table 3.1.1.1).

3.1.1.4 Dibenzo-p-Dioxins/Dibenzofurans (Dioxins/Furans, PCDDs/PCDFs)**Parameters****Congener Groups**

total tetrachlorodibenzo-p-dioxins (T4CDDs)
total pentachlorodibenzo-p-dioxins (P5CDDs)
total hexachlorodibenzo-p-dioxins (H6CDDs)
total heptachlorodibenzo-p-dioxins (H7CDDs)

2,3,7,8-Substituted Isomers

2,3,7,8-T4CDD
1,2,3,7,8-P5CDD
1,2,3,4,7,8-H6CDD
1,2,3,6,7,8-H6CDD
1,2,3,7,8,9-H6CDD

octachlorodibenzo-p-dioxin (O8CDD)	1,2,3,4,6,7,8-H7CDD
total tetrachlorodibenzofurans (T4CDFs)	1,2,3,4,6,7,8,9-O8CDD
total pentachlorodibenzofurans (P5CDFs)	2,3,7,8-T4CDF
total hexachlorodibenzofurans (H6CDFs)	1,2,3,7,8-P5CDF
total heptachlorodibenzofurans (H7CDFs)	2,3,4,7,8-P5CDF
	1,2,3,4,7,8-H6CDF
	1,2,3,6,7,8-H6CDF
	1,2,3,7,8,9-H6CDF
	2,3,4,6,7,8-H6CDF
	1,2,3,4,6,7,8-H7CDF
	1,2,3,4,7,8,9-H7CDF
octachlorodibenzo-p-furan (O8CDF)	1,2,3,4,6,7,8,9-O8CDF

Table 3.1.1.4

METHOD REFERENCE SOURCE	SOIL & SEDIMENT	GROUND WATER
USEPA	SW-846, Method 3545A SW-846, Method 3546 SW-846, Method 8290A Method 1613B	Method 1613B SW-846, Method 8290A
Environment Canada	EPSI/RM/19	EPSI/RM/19
MOE-LaSB	E3418	E3418

Method principle

This analytical method is used to determine the concentrations of dioxins (PCDDs) and furans (PCDFs) in a variety of matrices using isotope dilution with high resolution mass spectrometric detection (HRMS).

Solid samples are dried at ≤ 40 °C, ground, and homogenized. All samples are fortified prior to sample extraction, digestion, or elution, with known amounts of [$^{13}\text{C}_{12}$] isotopically labelled PCDDs and PCDFs. All analytes are quantified using isotope dilution against labelled standards. Solid samples are extracted using Soxhlet, microwave or pressurized liquid extraction (PLE) with solvent (Table 3.1.1.4), followed by a three-stage chromatographic cleanup procedure to remove any potential chemical interference.

Aqueous samples are extracted with solvent (Table 3.1.1.4), followed by a two-stage chromatographic cleanup procedure to remove potential chemical interferences. Extracts are stable indefinitely. The final extracts are analyzed using high resolution gas chromatography-high resolution mass spectrometry (HRGC-HRMS).

Calculation of toxic equivalents (TEQ)

There are a total of 210 dioxins and furans. Only 17 are toxic (2,3,7,8-substituted congeners) and their toxicity is normalized to 2,3,7,8-TCDD (the most toxic). The TEQ is determined (as shown in the following example) by multiplying the concentration of each detected 2,3,7,8-substituted congener by its respective toxic equivalent factor (TEF) to determine its toxic equivalence (TEQ).

The TEFs in the following table are those provided by the World Health Organization (WHO), 2005, as amended from time to time. For any 2,3,7,8-substituted congeners that are not detected, half of the estimated detection limit (EDL) is multiplied by the TEF to determine the TEQ for that congener. This converts each of the congeners to 2,3,7,8-TCDD toxic equivalents. The sum of the 17 toxic equivalents gives the TEQ (toxic equivalent) for the sample normalized to 2,3,7,8-TCDD. The result in this example is 1.64 pg/L.

Table 3.1.1.4.B. TEQ Example

Compound	CAS Number	Conc. pg/L	EDL pg/L	TEF	TEQ /congener pg/L
2,3,7,8-TCDD	1746-01-6	ND	1.1	1	0.55
1,2,3,7,8-PeCDD	40321-76-4	ND	1	1	0.5
1,2,3,4,7,8-HxCDD	39227-28-6	ND	1.2	0.1	0.06
1,2,3,6,7,8-HxCDD	57653-85-7	ND	0.89	0.1	0.045
1,2,3,7,8,9-HxCDD	19408-74-3	ND	1	0.1	0.05
1,2,3,4,6,7,8-HpCDD	35822-46-9	ND	1.1	0.01	0.0055
OCDD	3268-87-9	3.4		0.0003	0.00102
2,3,7,8-TCDF	51207-31-9	ND	1	0.1	0.05
1,2,3,7,8-PeCDF	57117-41-6	ND	1	0.03	0.015
2,3,4,7,8-PeCDF	57117-31-4	ND	1	0.3	0.15
1,2,3,4,7,8-HxCDF	70648-26-9	ND	0.82	0.1	0.041
1,2,3,6,7,8-HxCDF	57117-44-9	ND	1.1	0.1	0.055
2,3,4,6,7,8-HxCDF	60851-34-5	ND	1.1	0.1	0.055
1,2,3,7,8,9-HxCDF	72918-21-9	ND	1.2	0.1	0.06
1,2,3,4,6,7,8-HpCDF	67562-39-4	ND	0.95	0.01	0.0048
1,2,3,4,7,8,9-HpCDF	5567-89-7	ND	1	0.01	0.005
OCDF	39001-02-0	1.8		0.0003	0.00054

TOTAL TEQ 2,3,7,8-TCDD (0.5 DL) = 1.64 pg/L (sum of the TE/congener for each compound listed above)

TEQ = toxic equivalents = sum of individual TEQ/congener EDL = estimated detection limit

TEF = toxic equivalency factor

Report format

Report the total concentration of isomers detected in each congener group (e.g., total TCDD, etc.) to two significant figures; as well indicate the total number of isomers detected within that group. For non-detected target species or congener groups with no detected isomers, report the detection limit to one significant figure. Congener numbers need not be reported unless requested. The source and year of the TEF values used to calculate the TEQ must be identified.

3.1.1.5 Organochlorine Pesticides (OCs)

Parameters (Synonym)

aldrin	endosulfan II (thiodan sulphate II) ²
chlordane, <i>alpha</i> - (<i>α</i> -chlordane) ¹	endrin
chlordane, <i>gamma</i> - (<i>γ</i> -chlordane) ¹	heptachlor
DDD ³	heptachlor epoxide
DDE ³	hexachlorobenzene
DDT ³	hexachlorobutadiene
dieldrin	hexachloroethane
hexachlorocyclohexane, <i>gamma</i> - (<i>γ</i> -HCH,	methoxychlor (DMDT)
lindane, <i>γ</i> -BHC*)	
endosulfan I (thiodan sulphate I) ²	

*erroneously known as benzene hexachloride (BHC)

¹the sum of *alpha*- and *gamma*-chlordane is compared to the standard

² the sum of endosulfan I and II is compared to the standard

³DDT standard applies to the total DDT (i.e., sum of the DDT isomers), the DDE standard applies to total DDE (i.e., sum of the DDE isomers), and the DDD standard applies to the total DDD (i.e., sum of the DDD isomers.)

Table 3.1.1.5

METHOD REFERENCE SOURCE	SOIL & SEDIMENT	GROUND WATER
USEPA	<p>Sample Preparation SW-846, Method 3540C SW-846, Method 3541 SW-846, Method 3545A SW-846, Method 3546 SW-846, Method 3550C SW-846-3570</p> <p>Sample Cleanup SW-846 Method 3610B SW-846, Method 3620C SW-846, Method 3630C SW-846, Method 3660B</p> <p>Analysis SW-846, Method 8081B SW-846, Method 8270D</p>	<p>Sample Preparation SW-846, Method 3510C SW-846, Method 3520C SW-846, Method 3535A</p> <p>Sample Cleanup SW-846, Method 3610B SW-846, Method 3620C SW-846, Method 3630C SW-846, Method 3660B</p> <p>Analysis SW-846, Method 8081B SW-846, Method 8270D</p>
Standard Methods		Method 6410B Method 6630B Method 6630C
MOE-LaSB	E3487	E3400

Method principle

Each soil sample is extracted in a solvent or solvent mix (Table 3.1.1.5) Extraction methods include using Soxhlet extraction or ultrasonic bath followed by vortex shaker. Alternatively, pressurized fluid extraction may be used for soil or sediment samples.

Each aqueous sample is extracted with a solvent or solvent mix (Table 3.1.1.5). After extraction, a number of cleanup techniques may be applied, depending on the sample matrix and the determinative analytical method.

Soil and ground water extracts can be kept for 40 days.

Samples are analyzed using dual-column gas chromatography with electron capture detector (GC-ECD) or gas chromatography-mass spectrometer (GC-MS). The GC-ECD is more sensitive than the GC-MS for highly chlorinated compounds but the ECD is nonspecific and more subject to interferences. Thus, sample cleanup is required and second column confirmation of target analyte is required. Alternatively, samples may be analyzed by comprehensive two-dimensional gas chromatography with electron capture detector (GCxGC-ECD).

3.1.1.6 Petroleum Hydrocarbons (PHCs)

Parameters

petroleum hydrocarbons (PHCs) (C₆–C₁₀ Fraction)
F1 (C₆ to C₁₀)

petroleum hydrocarbons (PHCs) (C₁₀–C₅₀ Fraction)
F2 (C₁₀ to C₁₆), F3 (C₁₆ to C₃₄), F4[#] (C₃₄ to C₅₀), F4G[#] (gravimetric)

[#]the larger result obtained for F4 and F4G is compared to the standard

Table 3.1.1.6

METHOD REFERENCE SOURCE	SOIL & SEDIMENT	GROUND WATER
CCME	<p>Sample Preparation and Analysis Reference Method for the Canada-wide Standard for Petroleum Hydrocarbons (CWS-PHC) in Soil – Tier 1 Method</p> <p>Reference Method for the Canada-wide Standard for Petroleum Hydrocarbons (CWS-PHC) in Soil – Tier 1 Method – Addendum 1</p>	<p>Analysis Reference Method for the Canada-wide Standard for Petroleum Hydrocarbons (CWS-PHC) in Soil – Tier 1 Method</p> <p>Reference Method for the Canada-wide Standard for Petroleum Hydrocarbons (CWS-PHC) in Soil – Tier 1 Method – Addendum 1</p>

USEPA		SW 846 Method 5030B (F1) SW 846 Method 5021A (F1) SW 846 Method 8015C (F1– F4)
MOE-LaSB		E3421

PHCs in soil and sediment

Note: The analysis of petroleum hydrocarbons (PHCs) must be in accordance with the Canadian Council of Ministers of the Environment (CCME) method¹, which is composed of both “prescriptive” and “performance based” elements. The method also contains mandatory chromatography performance elements. For BTEX and PHC (F1) pre-charging the soil sampling container with methanol preservative is an accepted deviation from the CCME method.

Method principle

Fraction F1 is determined by processing a field-preserved soil or sediment sample as received (approximately 5 g) within 14 days from sampling, then analyzing by purge & trap or headspace gas chromatography with a flame ionization detector (GC-FID). Hermetic samplers and freezing are additional sample handling options that require modified preparation techniques. See Section 3.1.1.10 (VOCs) for details. Recovered methanol extracts are analyzed within 40 days from extraction.

Fractions F2, F3, F4 are determined by extracting a minimum of 5 g dry weight soil sample with 50:50 hexane/acetone in a Soxhlet apparatus or equivalent. The solvent recovered from the extracted soil sample is back extracted with water to remove and/or minimize the acetone content in the organic extract. The organic extract is dried using sodium sulphate and treated either *in situ* or by column chromatography with silica gel to remove polar material (50:50 dichloromethane/hexane). Recovered solvent extracts are analyzed within 40 days from extraction. The extract is analyzed by GC-FID.

Moisture content is determined as described in Section 2.1.1 (2).

For F1 the sample is analyzed by gas chromatography with a 100% polydimethylsiloxane (DB-1 or equivalent) column and a flame ionization detector. All area counts are integrated from the beginning of the nC₆ peak to the apex of the nC₁₀ peak to give F1. Standards containing nC₆, nC₁₀ and toluene are run. Toluene is used as a calibration standard. The nC₆ and nC₁₀ response factors must be within 30% of the response factor for toluene.

For F2, F3, F4, the sample is analyzed by gas chromatography with a 100% polydimethylsiloxane (DB-1 or equivalent) column and a flame ionization detector. It must be demonstrated daily that the average response factors for nC₁₀, nC₁₆ and nC₃₄ must be within 10% and the response factor of nC₅₀ must be within 30% of the average response factor for nC₁₀, nC₁₆ and nC₃₄. The hydrocarbon concentrations are calculated in the following three ranges.

¹ Canadian Council of Ministers of the Environment (CCME) “Reference Method for the Canada-wide Standard for Petroleum Hydrocarbons (CWS-PHC) in Soil – Tier 1 Method.”

- F2 result, nC₁₀ to nC₁₆ hydrocarbons, is determined by integration of all area counts from the apex of the nC₁₀ peak to the apex of the nC₁₆ peak. The average response factor for nC₁₀, nC₁₆ and nC₃₄ hydrocarbons is used for primary calibration.
- F3 result, nC₁₆ to nC₃₄ hydrocarbons, is determined by integration of all area counts from the apex of the nC₁₆ peak to the apex of the nC₃₄ peak. The average response factor for nC₁₀, nC₁₆ and nC₃₄ hydrocarbons is used for primary calibration.
- F4 result, nC₃₄ to nC₅₀ hydrocarbons, is determined by integration of all area counts from the apex of the nC₃₄ peak to the apex of the nC₅₀ peak. The average response factor for nC₁₀, nC₁₆ and nC₃₄ hydrocarbons is used for primary calibration. The GC response factor of the nC₅₀ must be within 30% of the average response factor of the nC₁₀, nC₁₆ and nC₃₄ hydrocarbons. This result gives fraction F4 provided that the chromatogram descends to baseline by the retention time of nC₅₀.
- F4G, gravimetric analysis, is determined if the chromatogram does not return to baseline at or before C₅₀. A 5 g or greater soil sample is extracted with 50:50 hexane/acetone. The solvent is evaporated and the weight of residue determined. If the result is less than 50% of the applicable standard specified for the soil texture and proposed use, stop the analysis and report this result. If the result is higher than 50% of the applicable site condition standard, the sample is reconstituted in 50:50 dichloromethane/hexane, treated with silica gel one time only, re-evaporated and the weight of residue determined. Both the F4 (GC) result and the F4G (gravimetric) result are reported, but the greater result is reported as Fraction F4. and used for purposes of comparison to the applicable site condition standard in **O. Reg. 153/04**.

F2–F4 analysis of high organic carbon soils

Soils and sediment with high organic content, such as peat, may exceed the capacity of the silica gel to remove non petroleum hydrocarbons. Another aliquot of the extract may be treated with a greater weight of silica gel if required. Gas chromatography-mass spectrometric (GC-MS) analysis may also be used to identify non-petroleum hydrocarbons. The reference method also suggests comparison to background samples. See the Reference Method for the Canada-wide Standard for Petroleum Hydrocarbons (CWS-PHC) in Soil – Tier 1 Method for additional detail.

PHCs in ground water

Note: A national method has not been approved for water samples. However, analysis of petroleum hydrocarbons (PHCs) in ground water must be in accordance with the CCME soil method and all performance requirements of the CCME method (Table 3.1.1.6).

Method principle

Fraction F1 is determined by purging a volume of a ground water sample, then analyzing by gas chromatography with a flame ionization detector (GC-FID).

Fractions F2, F3 and F4 are determined by extraction with hexane. Recovered extracts may be kept for up to 40 days from extraction. The solvent recovered from the extracted sample is dried using sodium sulphate and can be treated either *in situ* or by column chromatography with silica gel to remove polar material (50:50 dichloromethane/hexane). The extract is then analyzed by GC-FID.

For F1, the sample is analyzed by gas chromatography with a 100% polydimethylsiloxane (DB-1 or equivalent) column and a flame ionization detector. All area counts are integrated from the beginning

of the nC₆ peak to the apex of the nC₁₀ peak to give F1. Standards containing nC₆, nC₁₀ and toluene are run. Toluene is used as the calibration standard. The nC₆ and nC₁₀ response factors must be within 30% of the response factor for toluene.

For F2, F3, F4, the sample is analyzed by gas chromatography with a 100% polydimethylsiloxane (DB-1 or equivalent) column and a flame ionization detector as shown in method principle for petroleum hydrocarbons in soil and sediment.

Calculations:

For F1 in soil and sediment, the result is corrected for the soil moisture extracted into the methanol. The total solvent/water volume (Vt) is calculated using the following equation:

$$\text{Final Volume (methanol + water) mL} = \text{methanol volume mL} + (\% \text{moisture}/100 \times \text{wet sample wt g})$$

The results of PHC analysis need not include either benzene/toluene/ethylbenzene/xylenes (BTEX) or polycyclic aromatic hydrocarbons (PAHs). If concentrations of BTEX and/or PAHs are determined, both corrected and uncorrected results must be reported as follows:

F1, F1_{-BTEX}

F2, F2_{-naphthalene}

F3, F3_{-PAH}²

F4, F4G

3.1.1.7 Polychlorinated Biphenyls (PCBs)

Parameters

Aroclor 1242

Aroclor 1248

Aroclor 1254

Aroclor 1260

polychlorinated biphenyls, total

² PAH = phenanthrene, benz[a]anthracene, benzo[b]fluoranthene, benzo[k]fluoranthene, benzo[a]pyrene, fluoranthene, dibenz[a,h]anthracene, indeno[1,2,3-c,d]pyrene, pyrene

Table 3.1.1.7

METHOD REFERENCE SOURCE	SOIL & SEDIMENT	GROUND WATER
USEPA	<p>Sample Preparation SW-846, Method 3540C SW-846, Method 3541 SW-846, Method 3545A SW-846, Method 3546 SW-846, Method 3550 SW-846, Method 3570</p> <p>Sample Cleanup SW-846, Method 3610B SW-846, Method 3620C SW-846, Method 3630C SW-846, Method 3640A SW-846, Method 3660B SW-846, Method 3665A</p> <p>Analysis SW-846, Method 8082A SW-846, Method 8270C</p>	<p>Sample Preparation SW-846, Method 3510C SW-846, Method 3520C SW-846, Method 3535A</p> <p>Sample Cleanup SW-846, Method 3610B SW-846, Method 3620C SW-846, Method 3630C SW-846, Method 3640A SW-846, Method 3660B SW-846, Method 3665A</p> <p>Analysis SW-846, Method 8082A SW-846, Method 8270C</p>
Standard Methods		Method 6630B
ASTM		Method D5175-91 (2003)
USGS	O-5129-95	
Environment Canada		EPS 1/RM/31E
MOE-LaSB	E3487	E3400

Method principle

An aliquot of a solid sample is extracted with a solvent or solvent mix (Table 3.1.1.7). Extracts may be kept for up to 40 days. The extract is cleaned up using an approved reference method (Table 3.1.1.7) technique.

After sample cleanup, the extract is analyzed by injecting an aliquot of the sample into a gas chromatograph with an electron capture detector (GC-ECD). Analysis may be performed using single or dual column. Alternatively, samples may be analyzed by comprehensive two-dimensional gas chromatography with electron capture detector (GCxGC-ECD).

Aqueous samples are extracted, then concentrated, reconstituted, and analyzed by GC-ECD. Typical extraction solvents are methylene chloride or methylene chloride:hexane.

Alternatively, gas chromatography-mass spectrometry (GC-MS) may be used, provided the reporting limits (RLs) in Table 4 can be achieved and the quantitation protocol described below is used.

PCB Identification and Quantitation:

Four Aroclors are quantitated in this procedure: 1242, 1248, 1254 and 1260. Each Aroclor contains a mixture of individual PCB congeners that form a distinctive recognizable pattern in the chromatogram. Identification is accomplished by comparing the sample chromatogram to reference chromatograms of the individual Aroclors. Retention times and relative intensities of at least three major peaks must match the reference spectrum within specified limits for positive identification. Acceptance limits are retention times ± 6 seconds and ratios within $\pm 20\%$ of the reference Aroclor.

If the sample contains a single Aroclor, compare the response of the major peaks in the identified Aroclor to the reference Aroclor chromatogram and calculate the concentration of each. The average of the major peak concentrations is the concentration of the Aroclor (after including appropriate dilution factors).

If more than one Aroclor is identified and quantified, "total PCB" is the sum of the identified and quantitated Aroclors.

If a mixture or extreme weathering is evident, such that individual Aroclor identification is not possible, quantitate using a total area sum (excluding any "non-PCB" peaks) and comparing the sum against the Aroclor or Aroclor mixture that most closely resembles the sample. In this case "total PCB" is the result obtained using the total area sum.

3.1.1.8 Polycyclic Aromatic Hydrocarbons (PAHs) (may be analyzed with ABNs)

Parameters (Synonym)

acenaphthene	dibenz[a,h]anthracene
acenaphthylene	fluoranthene
anthracene	fluorene
benz[a]anthracene	indeno[1,2,3-cd]pyrene
benzo[a]pyrene (B[a]P)*	methylnaphthalene, 2-(1-)#
benzo[b]fluoranthene	naphthalene
benzo[g,h,i]perylene	phenanthrene
benzo[k]fluoranthene	pyrene
chrysene	

#the sum of 1- and 2-methylnaphthalene is compared to the standard

*benzo[a]pyrene (B[a]P) may be analyzed on filtered and unfiltered samples

Table 3.1.1.8

METHOD REFERENCE SOURCE	SOIL & SEDIMENT	GROUND WATER
USEPA	Sample Preparation SW-846, Method 3540C SW-846, Method 3541SW-846 Method 3546 SW-846, Method 3550C SW-846, Method 3570	Sample Preparation SW-846, Method 3510C SW-846, Method 3520C SW-846, Method 3535 SW-846, Method 3611B

	Sample Cleanup SW-846, Method 3610B SW-846, Method 3630C Analysis SW-846, Method 8270D	Analysis SW-846, Method 8270D
Standard Methods		Method 6440C
MOE-LaSB	E3425	E3480

Method principle

Soil samples fortified with deuterium-labelled surrogates are extracted using a solvent or solvent mix (Table 3.1.1.8). Cleanup of the extract is optional.

Aqueous samples, fortified with surrogates, are extracted with solvent (Table 3.1.1.8). If only PAHs are being determined, extraction may be at neutral or basic pH.

Extracts may be kept for up to 40 days. Duplicate samples may be submitted for filtration prior to B[a]P analysis. These samples are filtered using solvent rinsed glass wool or an equivalent filter into a new bottle prior to extraction. The sample extract is concentrated and then analyzed by means of gas chromatography-mass spectrometry (GC-MS), with or without using selected ion monitoring (SIM) mode.

See Section 3.1.1.1 (ABNs) for additional detail.

3.1.1.9 Trihalomethanes (THMs)*

Parameters (Synonyms)

bromodichloromethane (dichlorobromomethane)

bromoform (tribromomethane)

dibromochloromethane (chlorodibromomethane)

*may also be determined with ABNs or VOCs

Note that the above list of compounds are commonly detected as a result of chlorination of drinking water and, therefore, are included as a separate group from the volatile organic compounds (VOCs). The method principle for THMs is identical to the VOCs as outlined in Section 3.1.1.10 and Table 3.1.1.10.

3.1.1.10 Volatile Organic Compounds I (VOCs)

Parameters (Synonyms)

acetone	dichloropropene, <i>trans</i> -1,3-*
Bromomethane [#] (methyl bromide)	(dichloropropylene)
carbon tetrachloride (tetrachloromethane)	ethylene dibromide (dibromoethane, 1,2-)
chlorobenzene	hexane, n-**
chloroform (trichloromethane)	methyl ethyl ketone (MEK)
dichlorobenzene, 1,2-	methyl isobutyl ketone (MIBK)
dichlorobenzene, 1,3-	methyl <i>tert</i> -butyl ether (MTBE)
dichlorobenzene, 1,4-	methylene chloride (dichloromethane)
dichlorodifluoromethane	styrene
dichloroethane, 1,1-	tetrachloroethylene (tetrachloroethene, perchloroethylene)
dichloroethane, 1,2-	tetrachloroethane, 1,1,1,2-
dichloroethylene, 1,1- (dichloroethene)	tetrachloroethane, 1,1,2,2-
dichloroethylene, <i>cis</i> -1,2- (dichloroethene)	trichloroethane, 1,1,1-
dichloroethylene, <i>trans</i> -1,2- (dichloroethene)	trichloroethane, 1,1,2-
dichloropropane, 1,2-	trichloroethene (trichloroethylene, TCE)
dichloropropene, <i>cis</i> -1,3-*	trichlorofluoromethane
	vinyl chloride (chloroethene)

*the sum of *cis*- and *trans*-dichloropropene is compared to the standard

[#]methanol-preserved samples may elevate the detection limit for bromomethane; a separate bisulphate-preserved sample or hermetically sealed sample may be submitted at the time of sampling if bromomethane is a chemical of concern.

**may also be determined with BTEX

Table 3.1.1.10

METHOD REFERENCE SOURCE	SOIL & SEDIMENT	GROUND WATER
USEPA	<p>Sample Introduction SW-846, Method 5021A SW-846, Method 5035A Reference Method for the Canada-wide Standard for Petroleum Hydrocarbons (CWS-PHC) in Soil – Tier 1 Method</p> <p>Analysis SW-846, Method 8260C</p>	<p>Sample Introduction SW-846, Method 5000 SW-846, Method 5030B</p> <p>Analysis SW-846, Method 8260C EPA Method 624</p>
Standard Methods		Method 6200B
MOE-LaSB	E3254 E3490	E3132

VOCs in Soil and Sediment

Method principle

Volatile organic compounds (VOCs) present in field-preserved soil and sediment samples (approximately 5 g) are processed in the laboratory as received (Table 3.1.1.10) within 14 days from sampling. If required for bromomethane, duplicate samples preserved with aqueous sodium bisulphate can be analyzed as received.

Alternatively, unpreserved samples collected in hermetic sampling devices are extracted in the laboratory with methanol within 48 hours of sampling. Alternatively, to achieve a longer hold time, hermetic samples may be frozen within 48 hours of sampling as per ASTM method D6418 – 09; however, storage stability must be validated by the laboratory with no more than 10% losses.

Recovered methanol extracts are analyzed within 40 days from extraction.

Moisture content is determined as described in Section 2.1.1 (2).

Samples containing compounds exceeding the calibration range of the instrument are diluted by taking an aliquot of the extract diluted into volatile-free water and analyzed. Samples may be pre-screened by headspace gas chromatography-mass spectrometry (GC-MS) or other appropriate instrumentation to determine appropriate dilutions.

The volatile compounds present in the methanol or bisulphate solution are introduced by purge & trap or headspace into the gas chromatograph where they are separated by a capillary column and then detected by a mass spectrometer operating in either full scan or selected ion monitoring (SIM) mode.

Identification of target analytes is accomplished by comparing sample mass spectra with the mass spectra of analytical standards. Quantitation is accomplished by using the response of a major (quantitation) ion relative to an internal standard and a response factor generated from a calibration curve.

Calculations:

When reporting data based on a methanol extraction, concentrations must be corrected for the moisture extracted into the methanol (Table 3.1.1.10)

VOCs in ground water

Method principle

Aqueous samples are analyzed as received by purge & trap or headspace GC-MS (Table 3.1.1.10).

3.1.1.11 Volatile Organic Compounds II: Benzene, Ethylbenzene, Toluene, Xylenes (BTEX)*

Parameters (Synonyms)

benzene
ethylbenzene
Toluene (methylbenzene)
xylenes, total (o-xylene; m- & p-xylene)

*may also be determined with VOCs

Note that the above BTEX compounds (benzene, toluene, ethylbenzene, xylenes) are a subset of volatile organic compounds (VOCs), are often analyzed as a discrete analysis and, therefore, are included as a separate group from the VOCs. The method principle for BTEX is identical to the VOC I as outlined in Section 3.1.1.10 and Table 3.1.1.10.

3.1.2 INORGANIC CHEMICAL/PHYSICAL AND OTHER REGULATED PARAMETERS

3.1.2.1 Boron (B-HWS) (Hot Water Soluble)

Parameters

boron, hot water soluble

Table 3.1.2.1

METHOD REFERENCE SOURCE	SOIL & SEDIMENT	GROUND WATER
MOE-LaSB	<p>Sample Preparation Gupta, 1967, Soil Science 103: 424-428</p> <p>Analysis E3470</p>	Not applicable

Boron in soil or sediment

Method principle

A minimum 5 g portion of a dried, disaggregated (<2 mm) solid sample is extracted with 10 mL 0.01M calcium chloride (used to ensure a clear filtrate) through a 0.45µm filter. The sample is heated and must boil for five minutes followed by cooling and filtration. The sample is then analyzed using one of the spectrometric techniques listed in Table 3.1.2.1 or Table 3.1.2.8 (metals).

Note 5 g is the minimum size for a representative sample. Larger weights may be used but the 2:1 ratio v/w of aqueous calcium chloride to soil must be maintained.

Calculations and Reporting

All results are reported as µg/g dry weight.

3.1.2.2 Calcium and Magnesium

Parameters (Synonyms)

calcium (Ca)

magnesium (Mg)

Soil and sediment

Calcium and magnesium are determined for the calculation of sodium adsorption ratio (SAR) in soil and sediment samples. The method principle is outlined in Section 3.1.2.13 and Table 3.1.2.13 for SAR.

3.1.2.3 Chloride (Cl⁻) (water extractable)

Parameters

chloride (Cl⁻)

Table 3.1.2.3

METHOD REFERENCE SOURCE	SOIL & SEDIMENT	GROUND WATER
USEPA		SW-846, Method 6500 SW-846, Method 9056A SW-846, Method 9250 SW-846, Method 9251 SW-846, Method 9253 Method 300.0, Rev 2.1 Method 300.1, Rev 1.0
Standard Methods		Method 4110 B Method 4110 C Method 4500-Cl ⁻ C Method 4500-Cl ⁻ D Method 4500-Cl ⁻ E
MOE-LaSB	E3013	E3016

Chloride in soil or sediment

Method principle

A minimum 5 g portion of the previously dried, disaggregated (< 2mm) solid sample is extracted with 10 mL deionized water by shaking for a minimum of 30 minutes, then filtered and analyzed using ion chromatography or colourimetry. Note, 5 g is considered the minimum size for a representative sample. Larger weights may be used but the 2:1 ratio v/w of water to soil must be maintained.

In the colourimetric procedure, the chloride ions combine with mercuric thiocyanate to form an undissociated salt, mercuric chloride, and release thiocyanate ions which then complex ferric ions to produce a coloured solution. The absorbance of the coloured solution measured at the appropriate wavelength is proportional to the original concentration of chloride ion in the sample. The analysis is usually carried out using an automated continuous flow, flow injection or discrete analysis system.

Alternatively ion chromatography can be used. Ion chromatography is a form of liquid chromatography that uses ion-exchange resins to separate atomic or molecular ions based on their interaction with the resin.

Chloride in ground water

Method principle

Samples can be analyzed directly, or filtered in the laboratory prior to analysis using colourimetry or ion chromatography (Table 3.1.2.3).

3.1.2.4 Cyanide (CN⁻)

Parameters

Cyanide (CN⁻)

Table 3.1.2.4

METHOD REFERENCE SOURCE	SOIL & SEDIMENT	GROUND WATER
USEPA	Analysis SW-846, Method 9012B SW-846, Method 9014 SW-846, Method 9016	SW-846, Method 9014 Method 335.3 Method OIA-1677 SW-846, Method 9016
Standard Methods		Method 4500-CN-I Method 4500-CN-N Method 4500-CN-O
MOE-LaSB	E3015	E3015

Cyanide is defined as the simple and weakly dissociable cyanides that form hydrogen cyanide at pH 4.

Method principle

Soils and sediment

A minimum 10 g sample as received is extracted with 100 mL of 0.05N aqueous sodium hydroxide at a pH > 12. The sample is shaken for a minimum of six hours, followed by centrifuging and decanting. *Sodium hydroxide is used to ensure proper pH is maintained.* Larger weights may be used but the 10:1 ratio v/w of aqueous sodium hydroxide to soil must be maintained.

Ground water

Water samples are analyzed as received (Table 3.1.2.4). Particulates should not be included. Centrifugation or filtration in the laboratory may be required to remove particulate.

A portion of the aqueous sample or leachate is introduced directly to the autoanalyzer system from an autosampler. Cyanide is separated from water by auto acid distillation without UV oxidation at controlled pH and then analyzed colourimetrically or by means of a gas permeable membrane. Care must be taken in the analysis conditions to prevent thiocyanate interference. Off line distillation prior to analysis is an acceptable option.

The automated colourimetric method uses either barbituric plus isonicotinic acid, or pyridine plus barbituric acid, as colour reagent.

3.1.2.5 Electrical Conductivity

Table 3.1.2.5

METHOD REFERENCE SOURCE	SOIL & SEDIMENT	GROUND WATER
USEPA	SW-846, Method 9050A	n/a
MOE-LaSB	E3138	n/a

Electrical conductivity in soil and sediment

Method principle

A minimum 5 g sample of a previously dried, disaggregated (< 2 mm), sample is extracted with 10 mL pure water (20 mL for organic soils) by shaking for at least 30 minutes. The sample is then analyzed using a conductivity meter (Table 3.1.2.5). Note 5 g is considered the minimum size for a representative sample. Larger weights may be used but the 2:1 ratio v/w of water to soil must be maintained. Certain soil types may require a higher water:soil ratio in order to have sufficient liquid for measurement. This must be documented on the Certificate of Analysis or analytical report.

All aqueous solutions conduct electricity to various degrees. Conductivity is measured by an electronic meter or controller that applies an alternating voltage on the conductivity sensor and measures the resulting signal. The conductivity sensor consists of two or more electrodes of a certain area (A) separated by a predetermined distance (d). The sensor's cell constant (expressed in units of cm) is defined by:

$$K = d/A$$

The conductivity in $\mu\text{S}/\text{cm}$ is determined by multiplying the measured conductivity by the cell constant.

Report Format

The conductivity as measured in the extract is reported.

3.1.2.6 Hexavalent Chromium (Chromium VI, Cr (VI), Cr⁺⁶)

Parameters (Synonyms)

hexavalent chromium (chromium VI, Cr (VI), Cr⁺⁶)

Table 3.1.2.6

METHOD REFERENCE SOURCE	SOIL & SEDIMENT	GROUND WATER
USEPA	Sample Preparation SW-846, Method 3060A Analysis SW-846, Method 7196A SW-846, Method 7199	SW-846, Method 7196A SW-846, Method 7199 Method 218.6, Rev 3.3 ³ Method 1636
Standard Methods		Method 3500-Cr (2009)
ASTM		Method D5257-11
USGS	I-1232-85	I-1232-85

Hexavalent chromium in soil, sediment and ground water

Method principle

For soil and sediment samples, a minimum 2.5± 0.1 g sample as received is subjected to an alkaline digestion with continuous stirring prior to analysis (Table 3.1.2.6). The extract must be analyzed within seven days of extraction.

For the determination of dissolved hexavalent chromium, aqueous samples are field filtered and preserved with the ammonium sulfate buffer solution specified in USEPA Method 218.6 (revision 3.3, 1994) or Standard Methods 3500-Cr Chromium (2009) to a pH of 9.3 to 9.7 to achieve the 28 day³ holding time, or alternatively with sodium hydroxide with a holding time of 24 hours.

The most common analytical procedure is manual or automated colourimetry. The alkaline digestate (or base preserved aqueous sample) is acidified and treated with 1,5-diphenylcarbazide (DPC) which reacts with chromium VI to give a reddish-purple colour, the absorption of which is measured spectrophotometrically at a wavelength of 540 nm.

Alternatively, ion chromatography may be used for the analysis employing post column derivatization with DPC and measurement at 540 nm.

³ USEPA Federal Register Part III. March 12, 2007. 40 CFR Part 122, 136, et al. Guidelines Establishing Test Procedures for the Analysis of Pollutants Under the Clean Water Act; National Primary Drinking Water Regulations; and National Secondary Drinking Water Regulations; Analysis and Sampling Procedures; Final Rule, pages 11218, 11236, 11239 (footnote #20). Footnote 20 states: To achieve the 28-day holding time, use the ammonium sulfate buffer solution specified in EPA Method 218.6. The allowance in this footnote supersedes preservation and holding time requirements in the approved hexavalent chromium methods, unless this supersession would compromise the measurement, in which case requirements in the method must be followed.

3.1.2.7 Mercury (Hg)

Parameters

mercury (Hg)

Table 3.1.2.7

METHOD REFERENCE SOURCE	SOIL & SEDIMENT	GROUND WATER
USEPA	EPA 200.2 (Preparation) BCSALM (Preparation) SW-846, Method 3050B SW-846, Method 3051A SW-846, Method 7471B SW-846, Method 7474	EPA 200.2 (Preparation) SW-846, Method 7470A Method 245.1, Rev 3.0 Method 245.2 Method 245.7, Rev 2.0 Method 200.8, Rev 5.3 Method 1631E
Standard Methods		Method 3112 B
ASTM		Method D3223-02
USGS	I-16463-86	I-3462-85
MOE-LaSB	E3058 E3059	E3060

Method principle

Previously dried and ground (<0.355 mm) soil samples or aqueous samples are digested with a heated, strong, mixed acid solution to convert all forms of mercury to divalent mercury (Table 3.1.2.7). Excess oxidizing agents are removed by the addition of hydroxylamine. The divalent mercury is then reduced to elemental mercury, sparged from solution and analyzed in one of the following ways: manual or automated cold vapour atomic absorption spectrophotometry (CVAAS), or cold vapour atomic fluorescence spectrophotometry (CCAFS).

3.1.2.8 Methyl Mercury (Monomethyl Mercury, CH₃Hg⁺, MeHg⁺)

Parameters (Synonyms)

methyl mercury (monomethyl mercury, CH₃Hg⁺, MeHg⁺)

Table 3.1.2.8

METHOD REFERENCE SOURCE	SOIL & SEDIMENT	GROUND WATER
USEPA	SW-846, Method 3200 Method 1630	Method 1630

Method Principle

Soils: Extractable organomercury and inorganic mercury compounds are extracted from the soil matrix with acid using microwave or ultrasonic extraction. The organomercury compounds are separated using solid phase extraction or distillation and determined using a method for total mercury analysis (Table 3.1.2.8). Extracts/distillates must be analyzed within 48 hours of preparation.

Waters: Aqueous samples are acidified with hydrochloric acid forming organomercuric chloride (RHgCl) which is separated by distillation (Method 1630). The distillate is ethylated forming RHgEt. The volatile RHgEt complexes are purged onto a carbon trap, subsequently thermally desorbed, reduced to elemental mercury and detected by cold vapour atomic fluorescence spectrophotometry (CCAFS).

3.1.2.9 Metals

Parameters

barium (Ba)	molybdenum (Mo)
beryllium (Be)	nickel (Ni)
boron (B)	silver (Ag)
cadmium (Cd)	thallium (Tl)
chromium (Cr)	uranium (U)
cobalt (Co)	vanadium (V)
copper (Cu)	zinc (Zn)
lead (Pb)	

*strong acid extractable boron

Table 3.1.2.9

METHOD REFERENCE SOURCE	SOIL & SEDIMENT	GROUND WATER
USEPA	<p>Sample Preparation SW-846, Method 3050B SW-846, Method 3051A EPA 200.2</p> <p>Sample Analysis SW-846, Method 6010C SW-846, Method 6020A SW-846, Method 7000B SW-846, Method 7010 Method 200.2</p>	<p>Sample Preparation SW-846, Method 3005A SW-846, Method 3010A SW-846, Method 3015A SW-846, Method 3020A</p> <p>Sample Analysis SW-846, Method 6010C SW-846, Method 6020A SW-846, Method 7000B SW-846, Method 7010 Method 200.5, Rev 4.2 Method 200.7, Rev 4.4 Method 200.8, Rev 5.4 Method 200.9, Rev 2.2 Method 200.15, Rev 1.2</p>
Standard Methods		<p>Method 3111 B Method 3111 D Method 3113 B Method 3120 B Method 3125 B</p>

MOE-LaSB	E3470 E3075	E3474 E3386 E3094
----------	----------------	-------------------------

Method principle

Sample Preparation:

For soils, a previously dried, ground (<0.355 mm) sample is subjected to digestion with a heated, hydrochloric:nitric acid solution (Table 3.1.2.9). The digestate is separated from the soil residue and brought to volume with deionized water. This method provides environmentally available metals, not total metals.

Ground water samples requiring analysis for “dissolved” metals must be previously field filtered (0.45 micron) and field preserved to pH < 2. Unpreserved and unfiltered samples can be filtered and preserved at the laboratory provided that analysis not commence for at least 16 hours after preservation. This deviation must be noted on the C of A.

Analysis is performed with inductively coupled plasma-optical atomic emission spectroscopy (ICP-OES) or inductively coupled plasma-mass spectrometry (ICP-MS) or atomic absorption spectrophotometry (AAS) (Table 3.1.2.9)

The analytical standards must be matrix matched to the samples.

3.1.2.10 Metals, Hydride-Forming (As, Se and Sb)*

Parameters

antimony (Sb)

arsenic (As)

selenium (Se)

*may also be determined with metals by ICP-MS or ICP-OES

Table 3.1.2.10

METHOD REFERENCE SOURCE	SOIL & SEDIMENT	GROUND WATER
USEPA	<p>Sample Preparation SW-846, Method 3050B SW-846, Method 3051A</p> <p>Sample analysis SW-846, Method 6020A SW-846, Method 7010 SW-846, Method 7062 SW 846, Method 7762</p>	<p>Sample preparation SW-846, Method 3005A SW-846, Method 3010A SW-846, Method 3015A SW-846, Method 3020A</p> <p>Sample Analysis SW-846, Method 6020A SW-846, Method 7010 SW-846, Method 7061A SW-846, Method 7062 SW 846, Method 7762</p>

		Method 200.5, Rev 4.2 Method 200.8, Rev 5.4 Method 200.9, Rev 2.2
Standard Methods		Method 3113 B Method 3114 B Method 3114 C Method 3125 B
MOE-LaSB	E3245	E3474 E3089

Method principle

Sample Preparation:

Previously dried and ground (<0.355 mm) soil samples are extracted with a heated, mixed acid solution (Table 3.1.2.10). Filtered and preserved water samples may be acid digested or analyzed as received.

Analysis:

Samples are analyzed using graphite furnace atomic absorption spectrophotometry (GFAAS), inductively coupled plasma-mass spectrometry (ICP-MS) or hydride generation atomic absorption spectrophotometry (HGAAS). Alternatively, for elevated levels of arsenic, selenium and antimony, analysis can be performed with inductively coupled plasma-optical atomic emission spectroscopy (ICP-OES), and reported for each individual parameter five times above the ICP/OES method detection limit.

3.1.2.11 pH by Potentiometry

pH in soil, sediment

Table 3.1.2.11

METHOD REFERENCE SOURCE	SOIL & SEDIMENT	GROUND WATER
USEPA	SW-846, Method 9045D	n/a
MOE-LaSB	E3137	n/a

Method principle

A minimum 10 g portion of the sample, as received, is extracted with 20 mL of 0.01M calcium chloride solution by shaking for at least 30 minutes. The aqueous layer is separated from the soil by centrifuging, settling or decanting and then analyzed using a pH meter and electrode (Table 3.1.2.11). Note 10 g is considered the minimum size for a representative sample. Larger weights may be used but the 2:1 ratio v/w of aqueous calcium chloride to soil must be maintained.

The pH of a solution is defined as the negative logarithm of the hydrogen ion activity and in dilute solutions the activity is approximately equal to the concentration of the hydrogen ion. Thus

$$pH = -\log_{10} [H^+]$$

Since the activity of the hydrogen ion cannot be measured directly it is measured potentiometrically with a glass electrode in combination with a reference electrode

Report Format

The pH as measured in the extract is reported in pH units

3.1.2.12 Sodium (Na)

Parameters

sodium (Na)

Table 3.1.2.12

METHOD REFERENCE SOURCE	SOIL & SEDIMENT	GROUND WATER
USEPA	SW-846, Method 6010C SW-846, Method 6020A SW-846, Method 7000 B	Sample Preparation SW-846, Method 3005A SW-846, Method 3010A SW-846, Method 3015A SW-846, Method 3020A Sample Analysis SW-846, Method 6010C SW-846, Method 6020A SW-846, Method 7000B SW-846, Method 7010 Method 200.5, Rev 4.2 Method 200.7, Rev 4.4 Method 200.9, Rev 2.2 Method 200.15, Rev 1.2
Standard Methods		Method 3111 B Method 3111 D Method 3113 B Method 3120 B Method 3125 B
MOE-LaSB		E3474 E3386 E3094

Method principle

For soils and sediment samples, sodium (calcium and magnesium) is used to calculate the sodium adsorption ration (SAR). The method principle is outlined in Section 3.1.2.13 for SAR.

For sodium in ground water the method principle is outlined in Section 3.1.2.9 for metals.

3.1.2.13 Sodium Adsorption Ratio (SAR)

SAR in Soil and Sediment

Table 3.1.2.13

METHOD REFERENCE SOURCE	SOIL & SEDIMENT	GROUND WATER
USEPA	SW-846, Method 6010C SW-846, Method 6020A SW-846, Method 7000 B	

Method principle

A 5 g portion of previously dried, disaggregated sample (< 2 mm) is extracted with 10 mL deionized water by shaking for 30 minutes. For some soil types a higher water:soil ratio may be required to obtain sufficient liquid for measurement. The aqueous extract is separated from the solid, acidified and then analyzed using a spectrometric technique. Inductively coupled plasma-optical atomic emission spectroscopy (ICP-OES) is recommended, and alternatives are atomic emission spectroscopy (AAS) and inductively coupled plasma-mass spectrometry (ICP-MS) (Table 3.1.2.13).

Report format

The concentrations of sodium, calcium and magnesium are in units of milliequivalents per litre. SAR is determined from the equation below. Since SAR is a ratio, it is unitless.

$$SAR = \frac{[Na^+]}{\sqrt{\frac{1}{2}([Ca^{2+}] + [Mg^{2+}])}}$$

If any of the values are below the reporting detection limit (RDL), zero is used in the calculation for those parameters.

3.1.2.14 Fraction of Organic Carbon (FOC)

Fraction of Organic Carbon in Soil or Sediment

Table 3.1.2.14

METHOD REFERENCE SOURCE	SOIL & SEDIMENT	GROUND WATER
USEPA		n/a
ASTM	Method D2974-00 Method E1915-07	n/a
MOE-LaSB	E3142, E3012	n/a

Method principle

Fraction of organic carbon (FOC) in soil is a measure of the ratio of the organic carbon in the soil relative to the mass of sample ($g_{\text{carbon}}/g_{\text{soil}}$). Total organic carbon (TOC) is calculated as the

difference from analysis for total carbon (TC) and total inorganic carbon (TIC). The measurement of total carbon in soils and sediments requires the destruction of both carbonate minerals (primarily calcite and dolomite) as well as organic carbon.

Oxygen is purged through the system, as the samples are combusted, oxidizing carbon to carbon dioxide (CO₂). The CO₂ is collected, passed through two traps to remove moisture and dust, and then measured by an infrared detector (TC in mg/g carbon). Inorganic carbon (carbonate carbon) is determined by the measurement of CO₂ evolved by the reaction of carbonate with strong acid solution swept by purified nitrogen through a potassium iodide scrubber into the cathode compartment of a coulometer (Table 3.1.2.14). The evolved CO₂ is quantitatively absorbed by the cathode solution and converted to a strong acid causing the indicator colour to fade. Base is electrically generated to titrate the solution back to the starting point (TIC in mg/g carbon).

Alternatively, approved wet chemical reference methods can be used (Table 3.1.2.14). In these procedures, soil, after carbonate removal using acid, is treated with excess acidic dichromate, which reacts with the organic carbon, oxidizing it to CO₂. The residual dichromate is titrated with ferrous ammonium sulphate and TOC calculated by difference.

Results are reported in triplicate; separately and as an average.

SECTION 4: REPORTING

4.1 REQUIRED REPORTING LIMITS (RLs)

Reporting limit is the concentration at which a single analysis using the methods and matrices listed in this document will consistently detect target analytes when present. The RL must be equal to or greater than the method detection limit (MDL). The reporting limits listed in Table 4.1.1 apply in any situation where Table 4.1.2 does not apply.

The reporting detection limits listed in Table 4.1.2 apply when the applicable site condition standards with respect to which the sample submission is to be compared are derived from the document "Soil, Ground Water and Sediment Standards for Use Under Part XV.1 of the *Environmental Protection Act*, March 9, 2004". Note: Section 6 of that document applies to the reporting detection limits listed in Table 4.1.2.

TABLE 4.1.1 Required Reporting Limits

Chemical Name (Inorganics in bold)	CAS RN	Parameter Group	Soil Required RL µg/g (=mg/kg)	Ground Water Required RL µg/L
Acenaphthene	83-32-9	PAH or ABN	0.05	1
Acenaphthylene	208-96-8	PAH or ABN	0.05	1
Acetone	67-64-1	VOC	0.5	30
Aldrin	309-00-2	OC pesticides	0.05	0.01
Anthracene	120-12-7	PAH or ABN	0.05	0.1
Antimony	7440-36-0	hydride metals	1	0.5
Arsenic	7440-38-2	hydride metals	1	1
Barium	7440-39-3	metals	5	2
Benzene	71-43-2	VOC	0.02	0.5
Benz[a]anthracene	56-55-3	PAH or ABN	0.05	0.2
Benzo[a]pyrene	50-32-8	PAH or ABN	0.05	0.01 ^F
Benzo[b]fluoranthene	205-99-2	PAH or ABN	0.05	0.1
Benzo[ghi]perylene	191-24-2	PAH or ABN	0.1	0.2
Benzo[k]fluoranthene	207-08-9	PAH or ABN	0.05	0.1
Beryllium	7440-41-7	metals	2	0.5
Biphenyl, 1,1'-	92-52-4	ABN	0.05	0.5
Bis(2-chloroethyl)ether	111-44-4	ABN	0.5	5
Bis(2-chloroisopropyl)ether	39638-32-9	ABN	0.5	4
Bis(2-ethylhexyl)phthalate	117-81-7	ABN	5	10
Boron HWS (hot water soluble)	n/a	ORP	0.5	
Boron (Total)	7440-42-8	metals	5	10
Bromodichloromethane	75-27-4	THM or VOC	0.05	2
Bromoform	75-25-2	THM or VOC	0.05	5
Bromomethane	74-83-9	VOC	0.05	0.5
Cadmium	7440-43-9	metals	1	0.5
Carbon tetrachloride	56-23-5	VOC	0.05	0.2
Chlordane	57-74-9	OC pesticides	0.05	0.06
Chloride	16877-00-6	ORP	5	1000
Chloroaniline, p-	106-47-8	ABN	0.5	10
Chlorobenzene	108-90-7	VOC	0.05	0.5
Chloroform	67-66-3	VOC	0.05	1
Chlorophenol, 2-	95-57-8	CP or ABN	0.1	2

Chemical Name (Inorganics in bold)	CAS RN	Parameter Group	Soil Required RL µg/g (=mg/kg)	Ground Water Required RL µg/L
Chromium Total	7440-47-3	metals	5	10
Chromium VI	18540-29-9	ORP	0.2	10
Chrysene	218-01-9	PAH or ABN	0.05	0.1
Cobalt	7440-48-4	metals	2	1
Copper	7440-50-8	metals	5	5
Cyanide (CN-)	57-12-5	ORP	0.05	5
Dibenz[a,h]anthracene	53-70-3	PAH or ABN	0.1	0.2
Dibromochloromethane	124-48-1	THM or VOC	0.05	2
Dichlorobenzene, 1,2-	95-50-1	VOC	0.05	0.5
Dichlorobenzene, 1,3-	541-73-1	VOC	0.05	0.5
Dichlorobenzene, 1,4-	106-46-7	VOC	0.05	0.5
Dichlorobenzidine, 3,3'-	91-94-1	ABN	1	0.5
Dichlorodifluoromethane	75-71-8	VOC	0.05	2
DDD	72-54-8	OC pesticides	0.05	0.05
DDE	72-55-9	OC pesticides	0.05	0.01
DDT	50-29-3	OC pesticides	0.05	0.05
Dichloroethane, 1,1-	75-34-3	VOC	0.05	0.5
Dichloroethane, 1,2-	107-06-2	VOC	0.05	0.5
Dichloroethylene, 1,1-	75-35-4	VOC	0.05	0.5
Dichloroethylene, 1,2- <i>cis</i> -	156-59-2	VOC	0.05	0.5
Dichloroethylene, 1,2- <i>trans</i> -	156-60-5	VOC	0.05	0.5
Dichlorophenol, 2,4-	120-83-2	CP or ABN	0.1	20
Dichloropropane, 1,2-	78-87-5	VOC	0.05	0.5
Dichloropropene, 1,3- (<i>cis</i> - + <i>trans</i> -)	542-75-6	VOC	0.05	0.5
Dieldrin	60-57-1	OC pesticides	0.05	0.05
Diethyl phthalate	84-66-2	ABN	0.5	2
Dimethylphthalate	131-11-3	ABN	0.5	2
Dimethylphenol, 2,4-	105-67-9	ABN	0.2	10
Dinitrophenol, 2,4-	51-28-5	ABN	2	10
Dinitrotoluene, 2,4-(2,6-)	121-14-2	ABN	0.5	5
Dioxane, 1,4-	123-91-1	ABN or VOC	0.2	20
Dioxin/furan (TEQ)	n/a	dioxin/furan		
Electrical Conductivity (mS/cm)	n/a	ORP	0.005	
Endosulfan I/II	115-29-7/195-59-6/33213-65-9	OC pesticides	0.04	0.05
Endrin	72-20-8	OC pesticides	0.04	0.05
Ethylbenzene	100-41-4	VOC	0.05	0.5
Ethylene dibromide (dibromoethane, 1,2-)	106-93-4	VOC	0.05	0.2
Fluoranthene	206-44-0	PAH or ABN	0.05	0.4
Fluorene	86-73-7	PAH or ABN	0.05	0.5
Fraction Organic Carbon	n/a	ORP		
Heptachlor	76-44-8	OC pesticides	0.05	0.01
Heptachlor epoxide	1024-57-3	OC pesticides	0.05	0.01
Hexachlorobenzene	118-74-1	OC pesticides	0.01	0.01
Hexachlorobutadiene	87-68-3	OC pesticides	0.01	0.01
Hexachlorocyclohexane, <i>gamma</i> -	58-89-9	OC pesticides	0.01	0.01
Hexachloroethane	67-72-1	OC pesticides	0.01	0.01
Hexane, n-	110-54-3	VOC	0.05	5
Indeno[1,2,3-c,d]pyrene	193-39-5	PAH or ABN	0.1	0.2
Lead	7439-92-1	metals	10	1
Mercury	7439-97-6	ORP	0.1	0.1

Chemical Name (Inorganics in bold)	CAS RN	Parameter Group	Soil Required RL µg/g (=mg/kg)	Ground Water Required RL µg/L
Methoxychlor	72-43-5	OC pesticides	0.05	0.05
Methyl ethyl ketone	78-93-3	VOC	0.5	20
Methyl isobutyl ketone	108-10-1	VOC	0.5	20
Methyl mercury	22967-92-6	ORP		
Methyl <i>tert</i> -butyl ether (MTBE)	1634-04-4	VOC	0.05	2
Methylene chloride	75-09-2	VOC	0.05	5
Methylnaphthalene, 2- (1-)	91-57-6/90-12-0	PAH or ABN	0.05	2
Molybdenum	7439-98-7	metals	2	0.5
Naphthalene	91-20-3	PAH or ABN	0.05	2
Nickel	7440-02-0	metals	5	1
Nitrate	84145-82-4	ORP	5	100
Nitrite	14797-65-0	ORP	5	100
Nitrogen (total)	7727-37-9	ORP	500	250
Pentachlorophenol	87-86-5	CP or ABN	0.1	0.5
Petroleum hydrocarbons F1	n/a	PHC	10	25
Petroleum hydrocarbons F2	n/a	PHC	10	100
Petroleum hydrocarbons F3	n/a	PHC	50	500
Petroleum hydrocarbons F4 [#]	n/a	PHC	50	500
Phenanthrene	85-01-8	PAH or ABN	0.05	0.1
Phenol	108-95-2	ABN	0.5	1
Polychlorinated biphenyls (total)	1336-36-3	PCB	0.3	0.2
Pyrene	129-00-0	PAH or ABN	0.05	0.2
Selenium	7782-49-2	hydride metals	1	5
Silver	7440-22-4	metals	0.5	0.3
Sodium Adsorption Ratio	n/a	ORP		
Sodium	7440-23-5	metals and ORP	50	5000
Styrene	100-42-5	VOC	0.05	0.5
Tetrachloroethane, 1,1,1,2-	630-20-6	VOC	0.05	0.5
Tetrachloroethane, 1,1,2,2-	79-34-6	VOC	0.05	0.5
Tetrachloroethylene	127-18-4	VOC	0.05	0.5
Thallium	7440-28-0	metals	1	0.5
Toluene	108-88-3	VOC	0.2	0.5
Trichlorobenzene, 1,2,4-	120-82-1	ABN	0.05	0.5
Trichloroethane, 1,1,1-	71-55-6	VOC	0.05	0.5
Trichloroethane, 1,1,2-	79-00-5	VOC	0.05	0.5
Trichloroethylene	79-01-6	VOC	0.05	0.5
Trichlorofluoromethane	75-69-4	VOC	0.05	5
Trichlorophenol, 2,4,5-	95-95-4	CP or ABN	0.1	0.2
Trichlorophenol, 2,4,6-	88-06-2	CP or ABN	0.1	0.2
Uranium	7440-61-1	metals	1	2
Vanadium	7440-62-2	metals	10	0.5
Vinyl chloride	75-01-4	VOC	0.02	0.5
Xylene mixture	1330-20-7	VOC	0.05	0.5
Zinc	7440-66-6	metals	30	5

CAS RN = Chemical Abstracts Service Registry Number

n/a = not applicable or not available

^Ffiltration prior to analysis on a duplicate sample is permitted

ORP = other regulated parameters (listed in Section 1.2.3)

[#]the larger result obtained for F4 and F4G is compared to the RL

TABLE 4.1.2 PREVIOUS VALUES: 2004 Analytical Protocol APPENDIX B: Soil, Sediment and Water Standards and RLs

Full Depth Generic Site Condition Standards in a Potable Ground Water Condition and Coarse Textured Soils (soil standard for inorganics in this table apply only where soil pH is 5.0 to 9.0)						
Contaminant	Soil Standards (other than sediment) (µg/g)		Potable Ground Water Standards (µg/L)		Sediment Standards (µg/g)	
	Agricultural or Other Property Use	Required RL	All Types of Property Use	Required RL	All Types of Property Use	Required RL
ACENAPHTHENE	15	1.5	20	2.0	N/V	
ACENAPHTHYLENE	100	10	310	31	N/V	
ACETONE	3.5	N/D	3000	N/D	N/V	
ALDRIN	0.05	0.005	0.01	0.005	0.002	0.005 [#]
ANTHRACENE	28	2.8	12	1.2	0.22	0.02
ANTIMONY	13	1.3	6.0	0.60	N/V	
ARSENIC	20	2	25	2.5	6	1
BARIUM	750	75	1000	100	N/V	
BENZENE	0.24	0.05	5.0	0.5	N/V	
BENZO(a)ANTHRACENE	6.6	0.66	0.2	0.1	0.32	0.03
BENZO(a)PYRENE	1.2	0.12	0.01	0.01	0.37	0.04
BENZO(b)FLUORANTHENE	12	1.2	0.2	0.05	N/V	
BENZO(g,h,i)PERYLENE	40	4	0.2	0.10	0.17	0.02
BENZO(k)FLUORANTHENE	12	1.2	0.2	0.05	0.24	0.02
BERYLLIUM	1.2	1.9 [#]	4.0	0.5	N/V	
BIPHENYL, 1,1-	0.89	N/D	350	35	N/V	
BIS(2-CHLOROETHYL)ETHER	0.66	N/D	4.4	2.5	N/V	
BIS(2-CHLOROISOPROPYL)ETHER	0.66	N/D	2.2	1.0	N/V	
BIS(2-EHYLHEXYL)PHTHALATE	100	N/D	6.0	5.0	N/V	
BORON (AVAILABLE)	1.5 ⁺	0.15	5000	500	N/V	
BROMODICHLOROMETHANE	0.12	0.05	5.0	0.50	N/V	
BROMOFORM	0.11	0.05	5.0	0.50	N/V	
BROMOMETHANE	0.061	N/D	3.7	0.5	N/V	
CADMIUM	3.0	1	5.0	0.5	0.6	1
CARBON TETRACHLORIDE	0.10	0.05	5.0	0.5	N/V	
CHLORDANE	0.29	0.029	0.04	0.03	0.007	0.01 [#]
CHLOROANILINE, p-	1.3	N/D	28	N/D	N/V	
CHLOROBENZENE	2.4	0.24	30	3.0	N/V	
CHLOROFORM	0.13	0.05	5.0	0.50	N/V	
CHLOROPHENOL, 2-	0.1	N/D	0.3	0.3	N/V	

CHROMIUM (TOTAL)	750	75	50	5.0	26	5
CHROMIUM (VI)	8.0	(2.0) N/D	50	10	N/V	
CHRYSENE	12	1.2	0.5	0.05	0.34	0.03
COBALT	40	4	100	10	50b	4
COPPER	150	15	23	5	16	5
CYANIDE (FREE)	100	10	52	20	0.1b	0.05
DIBENZO(a,h)ANTHRACENE	1.2	0.12	0.2	0.1	0.06	0.01
DIBROMOCHLOROMETHANE	0.09	0.05	5.0	0.50	N/V	
DICHLOROENZENE, 1,2- (o-DCB)	0.88	0.088	3.0	0.5	N/V	
DICHLOROENZENE, 1,3- (m-DCB)	30	3	630	63	N/V	
DICHLOROENZENE, 1,4- (p-DCB)	0.32	0.05	1.0	0.5	N/V	
DICHLOROENZIDINE, 3,3'-	1.3	N/D	83	N/D	N/V	
DDD	2.2	0.22	6.0	0.60	0.008	0.025
DDE	1.6	0.16	20	2.0	0.005	0.005
DDT (total)	1.6	0.16	0.05	0.05	0.007	0.025 [#]
DICHLOROETHANE, 1,1-	3	0.3	70	7.0	N/V	
DICHLOROETHANE, 1,2-	0.022	N/D	5.0	0.5	N/V	
DICHLOROETHYLENE, 1,1-	0.0024	N/D	0.66	0.5	N/V	
DICHLOROETHYLENE, CIS-1,2-	2.3	0.23	70	7.0	N/V	
DICHLOROETHYLENE, TRANS-1,2-	4.1	0.41	100	10	N/V	
DICHLOROPHENOL, 2,4-	0.3	N/D	0.3	0.3	N/V	
DICHLOROPROPANE, 1,2-	0.019	N/D	5.0	0.50	N/V	
DICHLOROPROPENE, 1,3-	0.0066	N/D	1.4	0.5	N/V	
DIELDRIN	0.05	0.01	0.02	0.01	0.002	0.01 [#]
DIETHYL PHTHALATE	0.71	N/D	30	N/D	N/V	
DIMETHYL PHTHALATE	0.7	N/D	30	N/D	N/V	
DIMETHYLPHENOL, 2,4-	0.94	N/D	140	14	N/V	
DINITROPHENOL, 2,4-	0.2	N/D	42	100 [#]	N/V	
DINITROTOLUENE, 2,4-	0.66	N/D	0.5	0.5	N/V	
DIOXIN/FURAN (ng TEQ/g soil) / (ng TEQ/L water)	0.01	0.005	0.015	0.0075	N/V	
ENDOSULFAN (Total)	0.18	0.06	0.35	0.06	N/V	
ENDRIN	0.05	0.025	0.05	0.025	0.003	0.025 [#]
ETHYLBENZENE	0.28	0.05	2.4	1.2	N/V	
ETHYLENE DIBROMIDE	0.0056	N/D	1.0	0.5	N/V	
FLUORANTHENE	40	4	130	13	0.75	0.08
FLUORENE	340	34	280	28	0.19	0.02
HEPTACHLOR	0.084	0.0084	0.04	0.005	N/V	

HEPTACHLOR EPOXIDE	0.06	0.006	3.0	0.30	0.005c	0.005
HEXACHLOROBENZENE	0.46	0.046	0.62	0.062	0.02	0.005
HEXACHLOROBUTADIENE	0.38	0.038	0.45	0.045	N/V	
HEXACHLOROCYCLOHEXANE, GAMMA	0.41	0.041	0.8	0.08	N/V	
HEXACHLOROETHANE	3.8	0.38	2.5	0.25	N/V	
INDENO(1,2,3-cd)PYRENE	12	1.2	0.2	0.1	0.2	0.02
LEAD	200	20	10	1.0	31	10
MERCURY	10	1	0.12	0.1	0.2	0.05
METHOXYCHLOR	4.0	0.4	0.3	0.03	N/V	
METHYL ETHYL KETONE	0.27	N/D	350	N/D	N/V	
METHYL ISOBUTYL KETONE	0.48	N/D	350	N/D	N/V	
METHYL MERCURY	6.8 ⁺⁺	N/D	0.12	N/D	N/V	
METHYL TERT BUTYL ETHER	5.7	N/D	700	N/D	N/V	
METHYLENE CHLORIDE	1.1	0.11	50	5.0	N/V	
METHYLNAPHTHALENE, 2-([*] 1-)	1.2	N/D	10	2.5	N/V	
MOLYBDENUM	5.0	2.5	7300	730	N/V	
NAPHTHALENE	4.6	0.46	21	2.1	N/V	
NICKEL	150	15	100	10	16	2.5
PENTACHLOROPHENOL	5.0	(0.5) N/D	30	6.0	N/V	
PETROLEUM HYDROCARBON F1	30	10	1000 ^{**}	100	N/V	
PETROLEUM HYDROCARBON F2	150	10	1000 ^{**}	100	N/V	
PETROLEUM HYDROCARBON F3	400	50	1000 ^{**}	500	N/V	
PETROLEUM HYDROCARBON F4	2800	50	1000 ^{**}	500	N/V	
PHENANTHRENE	40	4	63	6.3	0.56	0.05
PHENOL	40	N/D	4200	420	N/V	
POLYCHLORINATED BIPHENYLS	0.5	0.1	0.2	0.1	0.07	0.1 [#]
PYRENE	250	25	40	4.0	0.49	0.05
SELENIUM	2.0	1	10	5.0	N/V	
SILVER	20	2	1.2	0.5	0.5b	0.5
STYRENE	1.2	N/D	100	10	N/V	
TETRACHLOROETHANE, 1,1,1,2-	0.019	N/D	5.0	0.50	N/V	
TETRACHLOROETHANE, 1,1,2,2-	0.01	N/D	1.0	0.5	N/V	
TETRACHLOROETHYLENE	0.45	0.05	5.0	3.0	N/V	
THALLIUM	4.1	(1.0) N/D	2.0	0.5	N/V	
TOLUENE	2.1	0.21	24	2.4	N/V	
TRICHLOROENZENE, 1,2,4-	30	3	70	7.0	N/V	
TRICHLOROETHANE, 1,1,1-	26	2.6	200	20	N/V	

TRICHLOROETHANE, 1,1,2-	0.28	0.05	5.0	0.50	N/V	
TRICHLOROETHYLENE	1.1	0.11	50	5.0	N/V	
TRICHLOROPHENOL, 2,4,5-	3.2	(0.3) N/D	200	20	N/V	
TRICHLOROPHENOL 2,4,6-	0.66	(0.1) N/D	2.0	0.1	N/V	
VANADIUM	200	20	200	20	N/V	
VINYL CHLORIDE	0.003	N/D	0.5	0.2	N/V	
XYLENES (Total)	25	2.5	300	150	N/V	
ZINC	600	60	1100	110	120	25
ELECTRICAL CONDUCTIVITY (mS/cm)	0.7	0.07	N/A	0.005	N/A	
CHLORIDE	N/V	2.5	250 (mg/L)	25 (mg/L)	N/V	
NITRATE	N/V	N/D	10 (mg/L)	(1.0 mg/L) N/D	N/V	
NITRITE	N/V	N/D	1.0 (mg/L)	0.1(mg/L)	N/V	
SODIUM ADSORPTION RATIO (SAR)	5	(0.5) N/D	N/A	N/D	N/A	
SODIUM	N/V	25	200 (mg/L)	2 (mg/L)	N/V	

N/A = Not Applicable; not required for Potable Water or Sediment Standards

N/V = No Value listed for standard

N/D = Not Developed at LSB – MOE; laboratories are expected to achieve an MDL value less than the standard value provided

**No non-aqueous phase liquid detected; sum of F1 + F2 cannot exceed 1000 µg/L; sum of F3 + F4 cannot exceed 1000 µg/L

#For these parameters the RDL listed is higher than the quoted standard value. Where the RDL is higher than the standard value, analytical laboratories are required to use more sensitive techniques to achieve an RDL that is lower than the standard value provided

+Boron soil criterion based on Hot Water Extract

++Analysis for methyl mercury is only required when the total mercury criterion is exceeded

(*1-) 2-methyl naphthalene soil criterion is applicable to 1-methyl naphthalene with the provision that if both are detected in the soil, the sum of the two concentrations cannot exceed the soil criterion

4.2: REPORTING REQUIREMENTS

As required by **O. Reg. 153/04**, each laboratory must adhere to the reporting requirements of Section 47 (3)(5)(6).

Certificates of Analysis or analytical reports shall include at least the following:

Submitted site and client information, including sample identifiers, location, etc.

Time Markers:

- Date and time sampled (for each sample) (if provided)
- Date extracted or digested (for each sample/test)
- Date analyzed (for each sample/test)
- Date reported (for each sample/test)
- Comment that report supersedes previous reports when corrected reports are generated and differences identified

Data Reportables:

- Temperature of samples upon receipt including whether the samples are frozen
- Presence of custody seals and whether intact
- Any other issues impacting sample integrity
- Chain of custody for samples submitted to the laboratory or transhipped between laboratories

QC Reportables:

In order that the QP can properly assess the quality of the analytical data, all associated QC must be reported as follows:

- Laboratory duplicate analyses (including relative percent difference (RPD) for each parameter)
- Field/travel blank(s) (where applicable)
- Method blank(s)
- Laboratory control sample analyses
- Matrix spike analyses (where applicable) (including % recovery)
- Reference materials (where applicable)
- Surrogate recoveries (where applicable)
- QC reportables must include flags of QC exceedances

Analysis Reportables:

- Analytical data
- Reporting detection limits (RLs)
- Units
 - all data will be reported in the same units as in the regulation
 - all soil/sediment data to be reported as dry weight
- Data qualifiers (interference, dry weight, etc)
 - visible particulate in ground water samples must be noted in the Certificate of Analysis or analytical report
- If requested the analytical uncertainty associated with each

- measurement
- The title of the analytical method as described in the scope of accreditation including the reference method upon which the analytical method is based
- Remarks/Comments:
- Report any unusual behaviour noted in any step of the analytical process (such as sample inhomogeneity, headspace in a volatile organic compound (VOC), etc)
 - Any other regulatory required comments (e.g., CCME performance criteria compliance)
- Subcontract Analyses:
- Analysis conducted in third party laboratories, including sister laboratories, must be so indicated

4.3: SAMPLE DILUTION

When the concentration of one or more parameters in a multicomponent scan (or the single analyte in a one component test) exceeds the concentration of their respective highest calibration standard, sample dilution is required to more accurately quantify the parameter. When this is required, the reported detection limit (RL) for each target analyte must be adjusted (increased) in direct proportion to the dilution factor (DF).

The dilution factor is determined as follows:

$$DF = \frac{\text{Final Volume of Diluted Sample (mL)}}{\text{Sample Aliquot Volume (mL)}}$$

RL_d (the revised RL for the diluted sample) is determined as follows:

$$RL_d = DF \times RL$$

Situations that require reporting RL_d (as a result of dilution) may not satisfy RL reporting limits. Such increases in RL are acceptable, as long as all parameter results are at or below the applicable standard. Each laboratory must fully document all sample dilutions and appropriately qualify the data.

Analytical note: the post-dilution concentration of the highest reported parameter must be no less than 20% of the highest calibration standard in the method. This will avoid losses of precision and accuracy and unnecessarily high reporting limits for other parameters that did not require dilution.

In multicomponent analytical scans it is also permissible to report results of the undiluted sample for analytes within the calibration range (if review shows the data to be valid).

4.3.1 Elevated non-target analyte or matrix interferences resulting in RLs above the standard

Where matrix interferences or elevated target/non-target compounds are present, sample dilution is required. The dilution may result in some target analytes being reported with adjusted RLs (as per the above calculation), above the pertinent **O. Reg. 153/04** limit.

In these cases results are reported as “less than (<)” with a raised RDL, corresponding to the level of interference, which may result in an RL above the standard.

In these cases, the QP must review the analytes where the RLs exceed the standard and establish if the compounds are contaminants of concern identified in the Phase I Environmental Site Assessment (ESA). If they are contaminants of concern, consult with the laboratory. Additional effort or non-routine testing may be required to achieve the required RLs. Note, however, that in cases involving very “dirty” samples it may not be possible to accurately quantify some analytes at the regulatory standard.

SECTION 5: REQUIRED QUALITY ASSURANCE/QUALITY CONTROL (QA/QC)

This section provides the method specific quality assurance and quality control (QA/QC) requirements with respect to the sample processing, analysis and reporting of analytical data submitted for the purpose of **Ontario Regulation 153/04 Records of Site Condition – Part XV.1 of the *Environmental Protection Act* (O. Reg. 153/04)**.

5.1 ACCREDITATION:

Laboratories must be accredited by an internationally recognized accreditation body [e.g., Standards Council of Canada (SCC), or Canadian Association for Laboratory Accreditation (CALA)] in accordance with the International Standard ISO/IEC17025:2005 – General Requirements for the Competence of Testing and Calibration Laboratories (as amended from time to time).

5.2 METHOD VALIDATION:

All analytical methods providing data in support of the regulation must be properly validated and proven fit for purpose. The validation data must be available for inspection on request.

In the event that the technique employed has been in place for a significant period of time, ongoing method performance data may be used to demonstrate the method is valid and fit for purpose. Method blanks, laboratory duplicates, laboratory control samples and matrix spikes must all meet the performance criteria outlined in Tables 5-1 to 5-15. A minimum of 30 data points for each measure is required. Method detection limits (MDLs) and uncertainty data must be current and fit for purpose. The MDLs must be less than or equal to (\leq) the reporting limits (RLs) in Table 4.1.1 and 4.1.2 as determined as per section 5.3.

In addition, single blind proficiency testing (PT) samples (if available) must demonstrate ongoing acceptable performance.

At a minimum, initial validation must include the following items. Further guidance is provided in the “Ontario Ministry of the Environment Protocol for the Acceptance of Alternate Methods (PAAM) Version 1.4 January 2005”.

5.2.1 Demonstration of Acceptable Precision, Accuracy, Selectivity and Specificity:

Ground Water:

A minimum of two sets of eight aliquots of real or synthetic ground water (not containing the analytes of interest) are spiked with the analytes of interest in the routinely used sample containers. One set is spiked at 5–10 times the RL, the other set at or above midrange. The samples are carried through the entire analytical process.

Soil and Sediment:

A minimum of two different soil types must be analyzed. One must be a clay matrix, the other an organic matrix (containing more than three percent total organic carbon (> 3%TOC). Well homogenized composite samples are prepared and a minimum of five aliquots of each soil type is spiked with all the analytes of interest at 5–10 times the RL and at or above the midrange (20 samples total). If suitable reference materials are available, they may be used. The samples are carried through the entire analytical process.

Analysis and Acceptability Criteria:

If possible, the analyses should be split between two analysts. The precision and accuracy of the two analysts should be within a factor of 1.5 to demonstrate acceptable method ruggedness.

The relative percent difference (RPD) of the replicates, and matrix spike recovery are calculated. (Electrical conductivity and pH are exempt from requiring matrix spikes). The RPD and recoveries must meet the limits specified in Tables 5-1 to 5-15 as appropriate. If the native concentration of some analytes are greater than the matrix spike concentration for some parameters, the matrix spike limits don't apply. If certified reference materials (CRMs) are used, the acceptance limits associated with the CRM must be met.

5.3 METHOD DETECTION LIMITS:

Method detection limits (MDLs) must be determined for every regulated parameter analyzed (except pH). If more than one instrument is used for a test, MDLs must be established for each instrument. The MDLs are redetermined at a minimum every two years or whenever major changes are made to the method or instrument.

The MDLs must be less than or equal to (\leq) the required reporting limit (RL) for each parameter.

The MDLs must be determined using protocols established by the Ontario Ministry of the Environment and as described below:

Prepare a sample (usually reagent water or blank soil) fortified at a level 1–10 times the expected MDL for the analytes of interest. If the resultant calculated MDL is not within this range the determination must be repeated until the calculated MDL concentration is 1–10 times the spike concentration.

Take a minimum of eight aliquots of the sample and process each through the entire analytical method. If a blank measurement is required to calculate the measured level of analyte, obtain a separate blank measurement for each sample aliquot analyzed.

Calculate a result (x) for each sample or sample/blank pair.

Calculate the conventional standard deviation (S_1) of the replicate measurements as follows:

$$S_1 = \sqrt{\frac{\sum_{i=1}^n (x_i - \bar{x})^2}{n-1}}$$

where:

x_i = analytical results in the final method reporting units for the n replicate aliquots ($i = 1$ to n)

\bar{x} = average of the “ n ” replicate measurements

Outliers identified using Grubb’s test or Dixon’s Q test may be deleted, but a minimum of seven data points must remain.

An alternative is to use previously determined within run replicate analysis data and calculate the standard deviation (S_2) of the replicate measurements as follows.

$$S_2 = \sqrt{\frac{\sum_{i=1}^n (x_1 - x_2)_i^2}{2n}}$$

where:

x_1, x_2 = the two replicate results for each of the n replicate pairs (minimum $n = 40$)

Compute the MDL as follows:

$$MDL = t_{(n-1, \alpha=0.01)} S$$

where:

$t_{(n-1, \alpha=0.01)}$ = the Student's t distribution appropriate for a 99% confidence level given the degrees of freedom $n-1$

α = traditionally called the level of significance of the test and is considered to be a measure of the maximum probability of a Type I error for all distributions consistent with the null hypothesis.

S = standard deviation as determined above

Table of Student's t Values at the 99 Percent Confidence Level (1 sided test)

Number of Replicates	Degree of Freedom (n-1)	t (n-1)
7	6	3.143
8	7	2.998
9	8	2.896
10	9	2.821
11	10	2.764
16	15	2.602
21	20	2.528
26	25	2.485
31	30	2.457
∞	∞	2.326

5.3.1 Determination of MDL for Summed Parameters

For summed parameters such as total xylenes, the MDL is the square root of the sum of the squares of the individual component MDLs. For example, if the MDL for o-xylene is 0.02 and m/p-xylene is 0.03, the total xylene MDL will be 0.04:

$$MDL_{total\ xylenes} = \sqrt{(MDL_{o-xylene}^2) + MDL_{m/p-xylene}^2} = \sqrt{0.0004 + 0.0009} = 0.04$$

5.3.2 Determination of MDL for Sodium Adsorption Ratio (SAR)

For calculated parameters such as SAR, the MDL is determined as in every other case by multiplying the standard deviation of SAR (S_{SAR}) by three. The MDL of SAR (MDL_{SAR}) is estimated using the following four equations.

EQUATION 1:
$$\overline{SAR} = \frac{[\overline{Na}]}{\sqrt{\frac{I}{2}([\overline{Ca}] + [\overline{Mg}])}}$$

EQUATION 2:
$$R_{SAR} = \sqrt{\left[\left(\frac{S_{Na}}{\overline{Na}} \right)^2 \right] + \left[\frac{S_{Ca}^2}{\frac{I}{Ca^2} [\overline{Ca}^2 + \overline{Mg}^2]^2} \right] + \left[\frac{S_{Mg}^2}{\frac{I}{Mg^2} [\overline{Ca}^2 + \overline{Mg}^2]^2} \right]}$$

EQUATION 3:
$$S_{SAR} = R_{SAR} \times \overline{SAR}$$

EQUATION 4:
$$MDL_{SAR} = S_{SAR} \times 3$$

Where:

R_{SAR} = relative standard deviation of sodium adsorption ratio

S_{SAR} = standard deviation sodium adsorption ratio

S_{Na} = standard deviation of sodium ion

S_{Ca} = standard deviation of calcium ion

S_{Mg} = standard deviation of magnesium ion

\overline{SAR} = average value of sodium adsorption ratio

\overline{Na} = average value of sodium ion

\overline{Ca} = average value of calcium ion

\overline{Mg} = average value of magnesium ion

5.3.3 Calculation of Toxic Equivalence MDL

The toxic dioxin and furan isomer concentrations are used to calculate a toxic equivalence factor. The same principle described in 5.3.1 is used except the MDL is multiplied by the toxic equivalent factor (TEF), then squared. The TEF MDL is the square root of the sum of squares of the individual MDL times the TEF values. An example is given below.

Example: Calculation of Toxic Equivalence MDL

CONGENER	I-TEF ^a	MDL ^b	MDL x TEF	(MDL x TEF) ²
2378 TCDF	0.1	8.9	0.89	0.7921
12378PCDF	0.03	9.3	0.279	0.077841
23478PCDF	0.03	7.8	0.234	0.054756
123478 HxCDF	0.1	8.5	0.85	0.7225
123678 HxCDF	0.1	7.2	0.72	0.5184
234678 HxCDF	0.1	8.6	0.86	0.7396
123789 HxCDF	0.1	8.6	0.86	0.7396
1234678 HpCDF	0.01	12	0.12	0.0144
1234789 HpCDF	0.01	8.4	0.084	0.007056
OCDF	0.0003	15	0.0045	0.00002025
2378 TCDD	1	1.8	1.8	3.24
12378 PCDD	1	5.7	5.7	32.49
123478 HxCDD	0.1	3.7	0.37	0.1369
123678 HxCDD	0.1	6.2	0.62	0.3844
123789 HxCDD	0.1	23	2.3	5.29
1234678 HpCDD	0.01	9.5	0.095	0.009025
OCDD	0.0003	46	0.0138	0.00019044
				45.22 (SUM OF SQUARES)

MDL = 6.72 (SQRT – sum of squares)

The MDL for each of the 17 “toxic congeners” is determined from eight spiked samples. The standard deviation of the mean is multiplied by Student *t* value (3 if eight samples are analyzed).

The MDL for each of the 17 congeners is multiplied by its TEF to convert its value to equivalents of 2,3,7,8-TCDD.

These values are then squared and summed. The square root of the sum of squares is the MDL value for the 2,3,7,8-TCDD toxic equivalent quantity (TEQ).

^a
I-TEF = international toxic equivalent factor

^b
MDL = method detection limit for each individual congener

5.4 MEASUREMENT UNCERTAINTY

Uncertainty of measurement must be estimated and documented. There are several guidelines for the estimation of measurement uncertainty including those published by MOE, the International Organization for Standardization (ISO) and EURACHEM/Cooperation on International Traceability in Analytical Chemistry (CITAC). Every possible source of uncertainty must be evaluated, but only those exceeding one-third the largest source need to be included in estimating combined uncertainty. If method performance data are used to estimate uncertainty, studies should be conducted such that the number and range of effects, concentrations and matrices are varied to ensure that the conditions encountered under normal use of the method are represented.

Uncertainty of measurement must be estimated for all analytes and expressed as expanded uncertainty (U) at 95% confidence (k=2).

Measurand: The specific quantity subject to measurement, such as the concentration of an analyte.

Uncertainty: A non-negative parameter associated with the result of a measurement that characterizes the dispersion of the values that could reasonably be attributed to the measurand.

Uncertainty Component: Uncertainty of a result may rise from many possible sources. Each of the separate contributions to uncertainty is referred to as an uncertainty component.

Standard Uncertainty: Uncertainty components are evaluated by the appropriate method and each is expressed as a standard deviation and is referred to as a standard uncertainty.

Combined Standard Uncertainty: Standard uncertainty components are combined to produce an overall value of uncertainty known as the combined standard uncertainty. It is an estimated standard deviation equal to the positive square root of the sum of variances of all uncertainty components.

$$\mu_c = \sqrt{\sum \mu_i^2}$$

where:

μ_c = combined uncertainty of the result

μ_i = uncertainty of the individual component

Expanded Uncertainty: Expanded uncertainty (U) is obtained by multiplying the combined standard uncertainty by a coverage factor “k” to provide an interval within which the value of the measurand is believed to lie, with a specified level of confidence (e.g. 95%).

$$U = \mu_c \times k$$

where:

μ_c = combined uncertainty of the result

k = 2 (for 95% confidence level)

5.5 QUALITY CONTROL SAMPLES

The laboratory quality control (QC) samples routinely analyzed are method blanks, laboratory control samples, sample duplicates and matrix spikes. In addition, surrogates standards are employed for organic analysis. The acceptance limits for these QC samples are the metrics by which the quality of the associated laboratory data is demonstrated.

Note that all applicable QC samples as tabulated below must be analyzed *when sufficient sample is available*. Extractable organic tests on ground water require multiple containers and the containers cannot be subsampled. The QP is responsible for the submission of multiple samples. If multiple containers are not submitted matrix spike and duplicate samples cannot be provided.

Laboratory QC may be supplemented by various field QC samples such as blind field duplicates, field/travel blanks, field and equipment rinsate blanks and field/travel spikes. In general, acceptance limits for field QC are broader than laboratory QC.

As well as these QC samples, there are additional data quality related requirements associated with all analytical methods such as: number of calibration standards, calibration curve frequency and acceptance criteria, continuing calibration frequency and acceptance criteria and gas chromatography-mass spectrometry (GC-MS) tuning criteria.

The acceptance criteria specified in the reference method for these elements should be met. If there are deviations from the reference, they must be documented and valid reasons given.

Field/Travel Blank: is a sample of methanol preservative, reagent water or blank soil transported to and from the sampling location carried through the entire sampling and analytical process including all sample preparation steps.

Method Blank: is a sample of reagent water or blank soil carried through the entire analytical process including all sample preparation steps.

Laboratory Control Sample (LCS): is a sample of reagent water or blank soil spiked with the analytes of interest and carried through the entire analytical process including all sample preparation steps. In general, the LCS will be a second source standard and have a concentration near the midpoint of the calibration range.

$$\text{LCS Recovery} = \frac{[\text{measured concentration}]}{[\text{design concentration}]}$$

Sample Duplicate: is a second aliquot of a soil or water sample carried through the entire analytical process including all sample preparation steps. Note that since most water sample tests for extractable organic analytes consume the entire sample, duplicates are actually field duplicates, and are only possible if sufficient additional sample bottles are provided to the lab.

$$\text{Duplicate RPD} = \frac{([\text{sample}] - [\text{sample duplicate}])}{([\text{sample}] + [\text{sample duplicate}])/2} \times 100$$

For organic analysis, soils are analyzed as received. As such, duplicates are primarily a measure of sample homogeneity. Similarly, because water analyses are “whole bottle” tests, sample duplicates are field duplicates, subject to sampling variability. The duplicate acceptance limits contained in the following tables are based on homogeneous samples. If samples are visibly non-homogeneous,

repeat analysis is not required. Data are reported flagged as “exceedance due to sample heterogeneity”. For most inorganic tests, samples are taken from the original container and processed with the documented method that will be used for the real samples, so the above stipulations do not apply.

Matrix Spike is a second aliquot of a soil or water sample spiked, usually about mid range, with all analytes determined in the analysis and carried through the entire analytical process including all sample preparation steps. Note that for soil tests where the extraction is not intended to recover all the native analyte (chloride, cyanide, HWS boron), the spike is added post extraction. In general, the matrix spike will be a second source standard and have a concentration near the midpoint of the calibration range. Reference materials (RMs) may be used in place of matrix spikes where appropriate provided the matrix is similar to the samples and they contain all analytes in the test.

$$\text{Matrix Spike Recovery} = ([\text{spiked sample}] - [\text{unspiked sample}]) / [\text{spike}] \times 100$$

Because matrix spikes are also impacted by sample heterogeneity, the issues discussed in sample duplicates above may apply.

Surrogates are used for organic tests. All samples are spiked with compounds (usually deuterated analogs) representative of the analytes being determined but not found in environmental samples. The surrogates are spiked into the sample prior to any sample preparation steps and carried through the entire analytical process.

$$\text{Surrogate Recovery} = ([\text{measured concentration}] / [\text{theoretical concentration}]) \times 100$$

Internal Standards are used for some organic tests (i.e., ABNs, VOCs). A known amount of compound(s) (not present in the samples, but closely matching the chemical behavior of the compound(s) of interest) are added to every sample (including all QC samples) prior to analysis to quantitate by comparing the response of a major (quantitation) ion relative to an internal standard.

Continuing Calibration Verifications: (CCVs) are evaluated to determine whether the instrument was within acceptable calibration throughout period in which samples were analyzed (i.e., to verify that the initial calibration was applicable during the sample analyses). In general, failure of the CCV indicates that the initial calibration is no longer valid and should trigger recalibration and the reanalysis of the associated samples in the analytical sequence.

5.5.1 Acceptance Limits and Qualifiers

The pre-established ranges of acceptability tabulated below are in accord with the reference methods outlined in Section 3 of this document.

Multielement Scan Qualifier: as the number of analytes in a scan increases, so does the chance of a limit exceedance by random chance as opposed to a real method problem. Thus, in multielement scans, for the LCS and matrix spike, up to 10% of the analytes may exceed the quoted limits by up to 10% absolute and the spike is considered acceptable. For example, in a polycyclic aromatic hydrocarbon (PAH) scan of 17 analytes with matrix spike acceptance limits of 50–140%, 10% or one analyte may have a recovery outside of 50–140% by 10% absolute, i.e., a recovery of 40–150%.

Duplicate Qualifier: for duplicates as the measured result approaches the RL, the uncertainty associated with the value increases dramatically, thus duplicate acceptance limits apply only where the average of the two duplicates is greater than five times the RL.

Matrix Spike Qualifier: for matrix spikes, as the concentration of the native analyte increases, the uncertainty of the matrix spike recovery increases. (It is not possible to accurately quantitate a small difference between two large numbers). Thus, the matrix spike acceptance limits apply only when the concentration of the matrix spike is greater than or equal to the concentration of the native analyte.

Calculated Parameters: for calculated parameters, acceptance limits should reflect the uncertainty (μ_i) in each measurement (see section 5.4). This is especially important for parameters calculated by difference such as $F1_{-BTEX}$.

For example, in a sample with a summed BTEX concentration of 10 mg/L and an F1 concentration of 11 mg/L, each with an μ_i of 20% or about 2 mg/L, the μ_i of the $F1_{-BTEX}$ reported result is 1 mg/L \pm 2.8, a component uncertainty of 280%:

$$\mu_{F1_{-BTEX}} = \sqrt{(2^2 + 2^2)} = 2.8 = 280\%.$$

In this example the routine QC acceptance limits for BTEX and F1 obviously cannot apply. For additive parameters, the impact is much less. Thus, for parameters calculated by subtraction, QC acceptance limits are only applied to the individual components.

TABLE 5-1: Performance Criteria – Acid/Base Neutral Extractable Organic Compounds (ABNs), Chlorophenols (CPs), Polycyclic Aromatic Hydrocarbons (PAHs)

Required QA/QC Parameter	Data Quality Objective	Required Performance Standard	Included in Analytical Report/Certificate of Analysis
Method Blank	Laboratory contamination evaluation	<ul style="list-style-type: none"> Extracted with every batch or every 20 samples, whichever is more frequent Matrix-specific (e.g., water, soil) Laboratory filtered PAH water samples require a filter blank Target analytes should be less than the reporting limit (RL) 	<p>YES: If an analyte fails, corrective action is required. If it cannot be corrected (reinjection, reanalysis) the data are reported flagged</p>
Laboratory Control Sample (LCS)	Laboratory method accuracy without matrix effects	<ul style="list-style-type: none"> Every batch or every 20 samples whichever is more frequent Separate-source standard LCS percent recovery should be between 50–140% for all compounds except 30–130% for p-chloroaniline, 3,3-dichlorobenzidene, phenol, 2,4-dimethylphenol, DNP 	<p>YES: Reextract/reanalyze all associated samples, if possible. If not, report the data flagged for all failing analytes.</p>
Matrix Spike (or Reference Material as per Section 5.5)	Laboratory method accuracy with matrix effects, sample homogeneity	<ul style="list-style-type: none"> Extracted with every batch or every 20 samples, whichever is more frequent Laboratory filtered PAH water samples require a filter spike Separate-source standard Percent recoveries should be between 50–140% for all compounds except 30–130% for p-chloroaniline, 3,3-dichlorobenzidene, phenol, 2,4-dimethylphenol, DNP, soil and water 	<p>YES: If recovery is outside of specified limits, repeat if necessary. See 5.5 Matrix Spike. If not or if the repeat also fails report recovery and qualify the result</p>
Sample Duplicate	Sample homogeneity, laboratory method precision	<ul style="list-style-type: none"> Analyzed with every batch or every 20 samples, whichever is more frequent RPD should be $\leq 30\%$ for waters and $\leq 40\%$ for solids 	<p>YES: If RPD is within method specifications – no action required. If RPD fails, repeat analysis may be required, see 5.5 Sample Duplicate</p>
Surrogates	Accuracy in sample matrix	<ul style="list-style-type: none"> Surrogates should represent analytes of interest and be representative of compound class of target analytes (e.g., use deuterated PAH if analyzing for PAHs, use phenolic surrogates if analyzing for pentachlorophenol) Percent recoveries in soil and water should be between 50–140% for all compounds 	<p>YES: If recovery is outside of specified limits laboratory must report recovery and qualify the result</p>
Internal Standards (IS)	Laboratory accuracy, method accuracy in sample matrix	<ul style="list-style-type: none"> Minimum of 3 at retention times across GC run Area counts in samples must be between 50–200% of the area counts in associated continuing calibration standard (CCV) (Section 5.10 of 8260B). Retention times of internal standards should be within ± 6 seconds of retention times in the associated CCV 	<p>NO: If one or more internal standards are outside limits, reanalyze sample unless obvious interference present</p>
Quantitation	N/A	<ul style="list-style-type: none"> Quantitation must be based on IS calibration Laboratory must use the average response factor or regression curve generated from the associated initial calibration for quantitation of each analyte At least 1 qualifier ion (recommend 2) must be used and meet ratio requirements. See SW-846 for guidance. At low concentrations ratios may be expanded but the qualifier must be present for positive identification 	<p>NO</p>

TABLE 5-2: Performance Criteria – 1,4-Dioxane

Required QA/QC Parameter	Data Quality Objective	Required Performance Standard	Included in Analytical Report/Certificate of Analysis
Method Blanks	Laboratory contamination evaluation	<ul style="list-style-type: none"> Prepared with every batch or every 20 samples, whichever is more frequent Matrix-specific (e.g., water, soil) Target analytes should be less than the reporting limit (RL) 	YES: If an analyte fails, corrective action is required. If it cannot be corrected (reinjection, reanalysis) the data are reported flagged
Laboratory Control Sample (LCS)	Laboratory method accuracy without matrix effects	<ul style="list-style-type: none"> Every batch or every 20 samples whichever is more frequent Separate-source standard LCS percent recovery should be between 50–140% 	YES: Reprepare/reanalyze all associated samples, if possible. If not, report the data flagged for all failing analytes
Matrix Spike (or Reference Material as per Section 5.5)	Laboratory method accuracy with matrix effects, sample homogeneity	<ul style="list-style-type: none"> Extracted with every batch or every 20 samples, whichever is more frequent Separate-source standard Percent recoveries should be between 50–140% soil and water 	YES: If recovery is outside of specified limits, repeat if necessary. See 5.5 Matrix Spike. If not or if the repeat also fails report recovery and qualify the result
Sample Duplicate	Sample homogeneity, laboratory method precision	<ul style="list-style-type: none"> Analyzed with every batch or every 20 samples, whichever is more frequent RPD should be $\leq 30\%$ for waters and $\leq 50\%$ for solids 	YES: If RPD is within method specifications – no action required. If RPD fails, repeat analysis may be required, see 5.5 Sample Duplicate
Surrogates	Accuracy in sample matrix	<ul style="list-style-type: none"> Surrogates should represent analytes of interest and be representative of compound class of target analytes Percent recoveries in soil should be between 50–140%, soil and water Quantitation must be based on isotope dilution method Laboratory must use the average response factor or regression curve generated from the associated initial calibration for quantitation of each analyte At least 1 qualifier ion must be used and meet ratio requirements. See SW846 for guidance. At low concentrations ratios may be expanded but the qualifier must be present for positive identification 	YES: If recovery is outside of specified limits laboratory must report recovery and qualify the result
Quantitation	N/A		NO

TABLE 5-3: Performance Criteria – Dioxins/Furans

Required QA/QC Parameter	Data Quality Objective	Required Performance Standard	Included in Analytical Report/Certificate of Analysis
Method Blanks	Laboratory contamination evaluation	<ul style="list-style-type: none"> • Extracted with every batch or every 20 samples, whichever is more frequent • Matrix-specific (e.g., water, soil) • Target analytes should be less than the reporting limit (RL) 	YES: If an analyte fails, corrective action is required. If it cannot be corrected (reinjection, reanalysis) the data are reported flagged
Laboratory Control Sample (LCS)	Laboratory method accuracy without matrix effects	<ul style="list-style-type: none"> • Every batch or every 20 samples whichever is more frequent • Separate-source standard • LCS percent recovery should be between 50–150% for soil and water 	YES: Reextract/reanalyze all associated samples, if possible. If not, report the data flagged for all failing analytes
Matrix Spike (or Reference Material as per Section 5.5)	Laboratory method accuracy with matrix effects, sample homogeneity	<ul style="list-style-type: none"> • Extracted with every batch or every 20 samples, whichever is more frequent • Separate-source standard • Percent recoveries should be between 50–150% for soil and water 	YES: If recovery is outside of specified limits, repeat if necessary. See 5.5 Matrix Spike. If not or if the repeat also fails report recovery and qualify the result
Sample Duplicate	Sample homogeneity, laboratory method precision	<ul style="list-style-type: none"> • Analyzed with every batch or every 20 samples, whichever is more frequent • RPD should be ≤30% for waters and ≤40% for solids 	YES: If RPD is within method specifications – no action required. If RPD fails, repeat analysis may be required, see 5.5 Sample Duplicate
Surrogates	Accuracy in sample matrix	<ul style="list-style-type: none"> • Surrogates should represent analytes of interest and be representative of compound class of target analytes • Percent recoveries should be between 40–140% for soil and water 	YES: If recovery is outside of specified limits laboratory must report recovery and qualify the result

TABLE 5-4: Performance Criteria – Organochlorine (OC) Pesticides

Required QA/QC Parameter	Data Quality Objective	Required Performance Standard	Included in Analytical Report/Certificate of Analysis
Method Blanks	Laboratory contamination evaluation	<ul style="list-style-type: none"> • Extracted with every batch or every 20 samples, whichever is more frequent • Matrix-specific (e.g., water, soil) • Target analytes should be less than the reporting limit (RL) 	YES: If an analyte fails, corrective action is required. If it cannot be corrected (reinjection, reanalysis) the data are reported flagged
Laboratory Control Sample (LCS)	Laboratory method accuracy without matrix effects	<ul style="list-style-type: none"> • Every batch or every 20 samples whichever is more frequent • Separate-source standard • LCS percent recovery should be between 50–140% for soil and water 	YES; Reextract/reanalyze all associated samples, if possible. If not, report the data flagged for all failing analytes
Matrix Spike (or Reference Material as per Section 5.5)	Laboratory method accuracy with matrix effects, sample homogeneity	<ul style="list-style-type: none"> • Extracted with every batch or every 20 samples, whichever is more frequent • Separate-source standard • Percent recoveries should be between 50–140% for soil and water 	YES: If recovery is outside of specified limits, repeat if necessary. See 5.5 Matrix Spike. If not or if the repeat also fails report recovery and qualify the result
Sample Duplicate	Sample homogeneity, laboratory method precision	<ul style="list-style-type: none"> • Analyzed with every batch or every 20 samples, whichever is more frequent • RPD should be ≤30% for waters and ≤40% for solids 	YES: If RPD is within method specifications – no action required. If RPD fails, repeat analysis may be required, see 5.5 Sample Duplicate
Surrogates	Accuracy in sample matrix	<ul style="list-style-type: none"> • Surrogates should represent analytes of interest and be representative of compound class of target analytes • Percent recoveries should be between 50–140% for soil and water on both columns 	YES: If recovery is outside of specified limits laboratory must report recovery and qualify the result

TABLE 5-5: Performance Criteria – Polychlorinated Biphenyls (PCBs)

Required QA/QC Parameter	Data Quality Objective	Required Performance Standard	Included in Analytical Report/Certificate of Analysis
Method Blanks	Laboratory contamination evaluation	<ul style="list-style-type: none"> • Extracted with every batch or every 20 samples, whichever is more frequent • Matrix-specific (e.g., water, soil) • Target analytes should be less than the reporting limit (RL) 	YES: If an analyte fails, corrective action is required. If it cannot be corrected (reinjection, reanalysis) the data are reported flagged
Laboratory Control Sample (LCS)	Laboratory method accuracy without matrix effects	<ul style="list-style-type: none"> • Every batch or every 20 samples whichever is more frequent • Separate-source standard • LCS percent recovery should be between 60–140% for soil and water 	YES; Reextract/reanalyze all associated samples, if possible. If not, report the data flagged for all failing analytes
Matrix Spike (or Reference Material as per Section 5.5)	Laboratory method accuracy with matrix effects, sample homogeneity	<ul style="list-style-type: none"> • Extracted with every batch or every 20 samples, whichever is more frequent • Separate-source standard • Percent recoveries should be between 60–140% for soil and water 	YES: If recovery is outside of specified limits, repeat if necessary. See 5.5 Matrix Spike. If not or if the repeat also fails report recovery and qualify the result
Sample Duplicate	Sample homogeneity, laboratory method precision	<ul style="list-style-type: none"> • Analyzed with every batch or every 20 samples, whichever is more frequent • RPD should be ≤30% for waters and ≤40% for solids 	YES: If RPD is within method specifications – no action required. If RPD fails, repeat analysis may be required, see 5.5 Sample Duplicate
Surrogates	Accuracy in sample matrix	<ul style="list-style-type: none"> • Surrogates should represent analytes of interest and be representative of compound class of target analytes • Percent recoveries should be between 60–140% for soil and water 	YES: If recovery is outside of specified limits laboratory must report recovery and qualify the result

TABLE 5-6: Performance Criteria – Petroleum Hydrocarbons (PHCs)

Required QA/QC Parameter	Data Quality Objective	Required Performance Standard	Included in Analytical Report/Certificate of Analysis
Method Blanks	Laboratory contamination evaluation	<ul style="list-style-type: none"> • Extracted with every batch or every 20 samples, whichever is more frequent • Matrix-specific (e.g., water, soil) • Target analytes should be less than the reporting limit (RL) 	YES: If an analyte fails, corrective action is required. If it cannot be corrected (reinjection, reanalysis) the data are reported flagged
Laboratory Control Sample (LCS)	Laboratory method accuracy without matrix effects	<ul style="list-style-type: none"> • Every batch or every 20 samples whichever is more frequent • Separate-source standard • LCS percent recovery should be between 80–120% for soil and 60–140% for water 	YES: Reextract/reanalyze all associated samples, if possible. If not, report the data flagged for all failing analytes
Matrix Spike (or Reference Material as per Section 5.5)	Laboratory method accuracy with matrix effects, sample homogeneity	<ul style="list-style-type: none"> • Extracted with every batch or every 20 samples, whichever is more frequent • Separate-source standard prepared from gasoline or diesel/motor oil as appropriate • Percent recoveries should be between 60–140% soil and water 	YES: If recovery is outside of specified limits, repeat if necessary. See 5.5 Matrix Spike. If not or if the repeat also fails report recovery and qualify the result
Sample Duplicate	Sample homogeneity, laboratory method precision	<ul style="list-style-type: none"> • Analyzed with every batch or every 20 samples, whichever is more frequent • RPD should be ≤30% for waters and ≤30% for solids 	YES: If RPD is within method specifications – no action required. If RPD fails, repeat analysis may be required, see 5.5 Sample Duplicate
Surrogates	Accuracy in sample matrix	<ul style="list-style-type: none"> • Surrogates should represent analytes of interest and be representative of compound class of target analytes • Percent recoveries in soil should be between 60–140% for soil and water 	YES: If recovery is outside of specified limits laboratory must report recovery and qualify the result

TABLE 5-7: Performance Criteria – Volatile Organic Compounds (VOCs)

Required QA/QC Parameter	Data Quality Objective	Required Performance Standard	Included in Analytical Report/Certificate of Analysis
Field/Travel Blanks	Field contamination evaluation	<ul style="list-style-type: none"> Prepared with every batch of methanol-preserved soil samples Field/travel blank is reweighed at the lab and compared to tared weight to determine any loss of methanol Target analytes should be less than the reporting limit (RL). Note: acetone, methylene chloride, toluene and hexane are common laboratory artefacts. If any are > RDL the laboratory must comment on the impact on data quality. 	YES: If an analyte fails, corrective action is required. If it cannot be corrected (reinjection, reanalysis) the data are reported flagged
Method Blanks	Laboratory contamination evaluation	<ul style="list-style-type: none"> Prepared with every batch or every 20 samples, whichever is more frequent Matrix-specific (e.g., water, soil) Target analytes should be less than the reporting limit (RL). Note: acetone, methylene chloride, toluene and hexane are common laboratory artefacts. If any are > RDL the laboratory must comment on the impact on data quality. 	YES: If an analyte fails, corrective action is required. If it cannot be corrected (reinjection, reanalysis) the data are reported flagged
Laboratory Control Sample (LCS)	Laboratory method accuracy without matrix effects	<ul style="list-style-type: none"> Every batch or every 20 samples whichever is more frequent Separate-source standard LCS percent recovery should be between 50–140% for compounds that are gaseous at 20C and ketones, 60–130% for all others, soil and water 	YES: Reprep/reatalyze all associated samples, if possible. If not, report the data flagged for all failing analytes
Matrix Spike (or Reference Material as per Section 5.5)	Laboratory method accuracy with matrix effects, sample homogeneity	<ul style="list-style-type: none"> Extracted with every batch or every 20 samples, whichever is more frequent Separate-source standard Percent recoveries should be between 50–140% soil and water 	YES: If recovery is outside of specified limits, repeat if necessary. See 5.5 Matrix Spike. If not or if the repeat also fails report recovery and qualify the result
Sample Duplicate	Sample homogeneity, laboratory method precision	<ul style="list-style-type: none"> Analyzed with every batch or every 20 samples, whichever is more frequent RPD should be ≤30% for waters and ≤50% for solids 	YES: If RPD is within method specifications – no action required. If RPD fails, repeat analysis may be required, see 5.5 Sample Duplicate
Surrogates	Accuracy in sample matrix	<ul style="list-style-type: none"> Surrogates should represent analytes of interest and be representative of compound class of target analytes Percent recoveries in soil should be between 50–140%, soil and water 	YES: If recovery is outside of specified limits laboratory must report recovery and qualify the result
Internal Standards (IS)	Laboratory analytical accuracy and method accuracy in sample matrix	<ul style="list-style-type: none"> Minimum of 3 at retention times across GC run Area counts in samples should be between 50–200% of the area counts in associated continuing calibration standard (Section 5.10 of 8260B) Retention times of internal standards should be within ±6 seconds of retention times in associated continuing calibration standard 	NO: If one or more internal standards are outside limits, reanalyze sample unless obvious interference present
Quantitation	N/A	<ul style="list-style-type: none"> Quantitation must be based on IS calibration Laboratory must use the average response factor or regression curve generated from the associated initial calibration for quantitation of each analyte At least 1 qualifier ion (recommend 2) must be used and meet ratio requirements. See SW846 for guidance. At low concentrations ratios may be expanded but the qualifier must be present for positive identification 	NO

TABLE 5-8: Performance Criteria – Cyanide (CN⁻)

Required QA/QC Parameter	Data Quality Objective	Required Performance Standard	Included in Analytical Report/Certificate of Analysis
Method Blanks	Laboratory contamination evaluation	<ul style="list-style-type: none"> • Every batch or every 20 samples whichever is more frequent • Should be matrix-matched (same concentration of reagents as calibration and QC standards) and distilled with samples in batch • Target analytes should be less than the reporting limit (RL) 	YES: If the analyte fails, corrective action is required. If it cannot be corrected (reinjection, reanalysis) the data are reported flagged
Laboratory Control Sample (LCS)	Laboratory method accuracy without matrix effects	<ul style="list-style-type: none"> • Every batch or every 20 samples whichever is more frequent • Separate-source standard (soil or water) spiked post extraction for soil • LCS percent recovery should be between 80–120% 	YES: Reprepare/reanalyze all associated samples, if possible. If not, report the data flagged
Matrix Spike (or Reference Material as per Section 5.5)	Laboratory method accuracy with matrix effects, sample homogeneity	<ul style="list-style-type: none"> • Analyzed with every batch or every 20 samples, whichever is more frequent • Separate-source standard, spiked post extraction for soil • Percent recoveries should be between 70–130% soil and water 	YES: If recovery is outside of specified limits, repeat if possible. If not or if the repeat also fails report recovery and qualify the result
Sample Duplicate	Sample homogeneity, laboratory method precision	<ul style="list-style-type: none"> • Analyzed with every batch or every 20 samples, whichever is more frequent • Percent difference should be $\leq 20\%$ for water and $\leq 35\%$ for soils 	YES: If RPD is outside of specified limits, repeat if possible. If not or if the repeat also fails report recovery and qualify the result

TABLE 5-9: Performance Criteria – Electrical Conductivity (EC)

Required QA/QC Parameter	Data Quality Objective	Required Performance Standard	Included in Analytical Report/Certificate of Analysis
Method (Preparation) Blank	Laboratory contamination evaluation	<ul style="list-style-type: none"> • Every batch or every 20 samples whichever is more frequent • Target analytes should be less than the reporting limit (RL) 	YES: If the analyte fails, corrective action is required. If it cannot be corrected (reinjection, reanalysis) the data are reported flagged
Laboratory Control Sample (LCS)	Laboratory method accuracy without matrix effects	<ul style="list-style-type: none"> • Every batch or every 20 samples whichever is more frequent • Separate-source standard (soil or water) • LCS percent recovery should be between 90–110% 	YES: Reextract/reanalyze all associated samples, if possible. If not, report the data flagged
Matrix Spike (or Reference Material as per Section 5.5)	Laboratory method accuracy with matrix effects, sample homogeneity	N/A	NA
Sample Duplicate	Sample homogeneity, laboratory method precision	<ul style="list-style-type: none"> • Analyzed with every batch or every 20 samples, whichever is more frequent • RPD should be $\leq 10\%$ 	YES: If RPD is outside of specified limits, repeat if possible. If not or if the repeat also fails report recovery and qualify the result

TABLE 5-10: Performance Criteria – Fraction Organic Carbon (FOC), Chloride

Required QA/QC Parameter	Data Quality Objective	Required Performance Standard	Included in Analytical Report/Certificate of Analysis
Method (Preparation) Blank	Laboratory contamination evaluation	<ul style="list-style-type: none"> • Every batch or every 20 samples whichever is more frequent • Should be matrix-matched (same concentration of reagents as calibration and QC standards) and prepared with samples in batch • Target analytes should be less than the reporting limit (RL) 	YES: If the analyte fails, corrective action is required. If it cannot be corrected (re-injection, reanalysis) the data are reported flagged
Laboratory Control Sample (LCS)	Laboratory method accuracy without matrix effects	<ul style="list-style-type: none"> • Every batch or every 20 samples whichever is more frequent • Separate-source standard (soil or water) spiked post extraction for chloride in soil • LCS percent recovery should be between 70–130% 	YES: Reprepare/reanalyze all associated samples, if possible. If not, report the data flagged
Matrix Spike (or Reference Material as per Section 5.5)	Laboratory method accuracy with matrix effects, sample homogeneity	<ul style="list-style-type: none"> • Analyzed with every batch or every 20 samples, whichever is more frequent • Separate-source standard spiked post extraction for chloride in soil • Percent recoveries should be between 70–130% in soil and water 	YES: If recovery is outside of specified limits, repeat if possible. If not or if the repeat also fails report recovery and qualify the result
Sample Duplicate	Sample homogeneity, laboratory method precision	<ul style="list-style-type: none"> • For chloride samples: analyzed with every batch or every 20 samples, whichever is more frequent • For FOC samples: all samples are taken and analyzed in triplicate, therefore, additional duplicates are not required • RPD should be $\leq 20\%$ for water and $\leq 35\%$ for soils 	YES: If RPD is outside of specified limits, repeat if possible. If not or if the repeat also fails report recovery and qualify the result

TABLE 5-11: Performance Criteria – Hexavalent Chromium, Cr(VI)

Required QA/QC Parameter	Data Quality Objective	Required Performance Standard	Included in Analytical Report/Certificate of Analysis
Method (Preparation) Blank	Laboratory contamination evaluation	<ul style="list-style-type: none"> • Every batch or every 20 samples whichever is more frequent • Should be matrix-matched (same concentration of reagents as calibration and QC standards) and prepared with samples in batch • Target analytes should be less than the reporting limit (RL) 	YES: If the analyte fails, corrective action is required. If it cannot be corrected (reinjection, reanalysis) the data are reported flagged
Laboratory Control Sample (LCS)	Laboratory method accuracy without matrix effects	<ul style="list-style-type: none"> • Every batch or every 20 samples whichever is more frequent • Separate-source standard (soil or water) • LCS percent recovery should be between 80–120% 	YES; Reprep/reatalyze all associated samples, if possible. If not, report the data flagged
Matrix Spike (or Reference Material as per Section 5.5)	Laboratory method accuracy with matrix effects, sample homogeneity	<ul style="list-style-type: none"> • Analyzed with every batch or every 20 samples, whichever is more frequent • Separate-source standard • Percent recoveries should be between 70–130% soil* and water 	YES: If recovery is outside of specified limits, repeat if possible. If not or if the repeat also fails report recovery and qualify the result
Sample Duplicate	Sample homogeneity, laboratory method precision	<ul style="list-style-type: none"> • Analyzed with every batch or every 20 samples, whichever is more frequent • RPD should be ≤ 20% for water and ≤ 35% for soils 	YES: If RPD is outside of specified limits, repeat if possible. If not or if the repeat also fails report recovery and qualify the result

* Some soil samples may react with the Cr(VI) spike reducing it to Cr(III). These samples are highly unlikely to contain native hexavalent chromium. Thus a failed spike recovery does not invalidate a negative result on the native sample.

TABLE 5-12: Performance Criteria – Mercury

Required QA/QC Parameter	Data Quality Objective	Required Performance Standard	Included in Analytical Report/Certificate of Analysis
Method (Preparation) Blank	Laboratory contamination evaluation	<ul style="list-style-type: none"> • Every batch or every 20 samples whichever is more frequent • Should be matrix-matched (same concentration of reagents as calibration and QC standards) and prepared with samples in batch • Target analytes should be less than the reporting limit (RL) 	YES: If the analyte fails, corrective action is required. If it cannot be corrected (reinjection, reanalysis) the data are reported flagged
Laboratory Control Sample (LCS)	Laboratory method accuracy without matrix effects	<ul style="list-style-type: none"> • Every batch or every 20 samples whichever is more frequent • Separate-source standard (soil or water) • LCS percent recovery should be between 80–120% 	YES: Redigest/reanalyze all associated samples, if possible. If not, report the data flagged
Matrix Spike (or Reference Material as per Section 5.5)	Laboratory method accuracy with matrix effects, sample homogeneity	<ul style="list-style-type: none"> • Analyzed with every batch or every 20 samples, whichever is more frequent • Separate-source standard • Percent recoveries should be between 70–130% soil and water 	YES: If recovery is outside of specified limits, repeat if possible. If not or if the repeat also fails report recovery and qualify the result
Sample Duplicate	Sample homogeneity, laboratory method precision	<ul style="list-style-type: none"> • Analyzed with every batch or every 20 samples, whichever is more frequent • RPD should be $\leq 20\%$ for water and $\leq 30\%$ for soils 	YES: If RPD is outside of specified limits, repeat if possible. If not or if the repeat also fails report recovery and qualify the result

TABLE 5-13: Performance Criteria – Methyl Mercury

Required QA/QC Parameter	Data Quality Objective	Required Performance Standard	Included in Analytical Report/Certificate of Analysis
Method (Preparation) Blank	Laboratory contamination evaluation	<ul style="list-style-type: none"> • Every batch or every 20 samples whichever is more frequent • Should be matrix-matched (same concentration of reagents as calibration and QC standards) and prepared with samples in batch • Target analytes should be less than the reporting limit (RL) 	YES: If the analyte fails, corrective action is required. If it cannot be corrected (reinjection, reanalysis) the data are reported flagged
Laboratory Control Sample (LCS)	Laboratory method accuracy without matrix effects	<ul style="list-style-type: none"> • Every batch or every 20 samples whichever is more frequent • Separate-source standard (soil or water) • LCS percent recovery should be between 70–130% 	YES: Reprepare/reanalyze all associated samples, if possible. If not, report the data flagged
Matrix Spike (or Reference Material as per Section 5.5)	Laboratory method accuracy with matrix effects, sample homogeneity	<ul style="list-style-type: none"> • Analyzed with every batch or every 20 samples, whichever is more frequent • Separate-source standard • Percent recoveries should be between 60–140% soil and water 	YES: If recovery is outside of specified limits, repeat if possible. If not or if the repeat also fails report recovery and qualify the result
Sample Duplicate	Sample homogeneity, laboratory method precision	<ul style="list-style-type: none"> • Analyzed with every batch or every 20 samples, whichever is more frequent • RPD should be $\leq 30\%$ for water and $\leq 40\%$ for soils 	YES: If RPD is outside of specified limits, repeat if possible. If not or if the repeat also fails report recovery and qualify the result

TABLE 5-14: Performance Criteria – Boron, Hot Water Soluble (HWS); Calcium, Magnesium; Sodium; Metals (Including Hydride-Forming Metals)

Required QA/QC Parameter	Data Quality Objective	Required Performance Standard	Included in Analytical Report/Certificate of Analysis
Method (Preparation) Blank	Laboratory contamination evaluation	<ul style="list-style-type: none"> • Every batch or every 20 samples whichever is more frequent • Should be matrix-matched (same concentration of reagents as calibration and QC standards) and prepared with samples in batch • Target analytes should be less than the reporting limit (RL) 	YES: If the analyte fails, corrective action is required. If it cannot be corrected (reinjection, reanalysis) the data are reported flagged
Laboratory Control Sample (LCS)	Laboratory method accuracy without matrix effects	<ul style="list-style-type: none"> • Every batch or every 20 samples whichever is more frequent • Separate-source standard (soil or water), spiked post extraction for HWSB • LCS percent recovery should be between 80–120%. HWBS 70–130% 	YES: Reextract/reanalyze all associated samples, if possible. If not, report the data flagged
Matrix Spike (or Reference Material as per Section 5.5)	Laboratory method accuracy with matrix effects, sample homogeneity	<ul style="list-style-type: none"> • Analyzed with every batch or every 20 samples, whichever is more frequent • Separate-source standard, spiked post extraction for HWSB • Percent recoveries should be between 70–130% soil and water. HWSB 60–140% (soil) 	YES: If recovery is outside of specified limits, repeat if possible. If not or if the repeat also fails report recovery and qualify the result
Sample Duplicate	Sample homogeneity, laboratory method precision	<ul style="list-style-type: none"> • Analyzed with every batch or every 20 samples, whichever is more frequent • RPD should be $\leq 20\%$ for water and $\leq 30\%$ for soils. HWSB $\leq 40\%$ for soil 	YES: If RPD is outside of specified limits, repeat if possible. If not or if the repeat also fails report recovery and qualify the result

TABLE 5-15 Performance Criteria – pH in Soil

Required QA/QC Parameter	Data Quality Objective	Required Performance Standard	Included in Analytical Report/Certificate of Analysis
Method (Preparation) Blank	Laboratory contamination evaluation	N/A	N/A
Laboratory Control Sample (LCS)	Laboratory method accuracy without matrix effects	<ul style="list-style-type: none"> • Every batch or every 20 samples whichever is more frequent • Separate-source standard (soil or water) • LCS percent recovery should be between ± 0.2 pH units 	YES: reanalyze all associated samples, if possible. If not, report the data flagged
Matrix Spike (or Reference Material as per Section 5.5)	Laboratory method accuracy with matrix effects, sample homogeneity	N/A	N/A
Sample Duplicate	Sample homogeneity, laboratory method precision	<ul style="list-style-type: none"> • Analyzed with every batch or every 20 samples, whichever is more frequent • within 0.3 pH units 	YES: If RPD is outside of specified limits, repeat if possible. If not or if the repeat also fails report recovery and qualify the result

SECTION 6: REFERENCES

- American Society for Testing and Materials (ASTM), 2010 Annual Book of Standards
- American Society for Testing and Materials (ASTM) Method D6418 – 09: Standard Practice for Using the Disposable EnCore Sampler for Sampling and Storing Soil for Volatile Organic Analysis, Annual Book of Standards 2010, Volume 11.04
- Canadian Council of Ministers of the Environment (CCME). Reference Method for the Canada-wide Standard for Petroleum Hydrocarbons (CWS-PHC) in Soil – Tier 1 Method. 2001
- EURACHEM/Cooperation on International Traceability in Analytical Chemistry (CITAC), Quantifying Uncertainty in Analytical Measurement, (Second Edition, 2000)
- Gupta, UC 1967: A simplified method for determining hot water soluble boron in podzol soils, Soil Science 103: 424-428
- Handbook of Quality Assurance for the Analytical Chemistry Laboratory. Dux, James P. 1986. Van Nostrand Reinhold Company Inc
- International vocabulary of metrology – Basic and general concepts and associated terms; ISO/IEC Guide 99:2007 (VIM 2007)
- International Standard ISO/IEC 17025:2005 – General Requirements for the Competence of Testing and Calibration Laboratories
- Massachusetts Department of Environmental Protection Bureau of Waste Site Cleanup, Quality Assurance and Quality Control Requirements for SW-846 Method 7196A, Hexavalent Chromium by UV-Visible Spectrophotometry for the Massachusetts Contingency Plan (MCP)
- Quality Assurance and Quality Control Requirements for SW-846 Method 8081A, Chlorinated Pesticides by Gas Chromatography (GC) for the Massachusetts Contingency Plan (MCP)
- Quality Assurance and Quality Control Requirements for SW-846 Method 8082, Polychlorinated Biphenyls (PCBs) by Gas Chromatography (GC) for the Massachusetts Contingency Plan (MCP)
- Quality Assurance and Quality Control Requirements and Performance Standards for SW-846 Method 8151A, Chlorinated Herbicides by Gas Chromatography Using Methylation or Pentafluorobenzoylation Derivatization, for the Massachusetts Contingency Plan (MCP)
- Quality Assurance and Quality Control Requirements for SW-846 Method 8270C, Semivolatile Organic Compounds by Gas Chromatography/Mass Spectrometry (GC/MS) for the Massachusetts Contingency Plan (MCP)
- Quality Assurance and Quality Control Requirements and Performance Standards for SW-846 Method 9014, Total Cyanide and the MADEP Physiologically Available Cyanide (PAC) Protocol, for the Massachusetts Contingency Plan (MCP)
- Quality Assurance and Quality Control Requirements for SW-846 Method 8260B, Volatile Organic Compounds by Gas Chromatography/ Mass Spectrometry (GC/MS) for the Massachusetts Contingency Plan (MCP)
- Ontario Ministry of the Environment (MOE) Practices for the Collection and Handling of

Drinking Water Samples, Version 2.0, April 1, 2009

- Ontario Ministry of the Environment (MOE) Protocol for Analytical Methods Used in the Assessment of Properties under Part XV.1 of the *Environmental Protection Act*, March 9, 2004
- Ontario Ministry of the Environment (MOE) Protocol for the Acceptance of Alternate Methods (PAAM) Version 1.4 January 2005
- Ontario Ministry of the Environment, Protocol of Accepted Drinking Water Testing Methods, Version 2.0, May 2010
- Ontario Ministry of the Environment, Protocol for Sampling and Analysis of Industrial/Municipal Waste Water, Version 2.0, January 2011
- Ontario Ministry of the Environment (MOE), Laboratory Services Branch, Quality Management Section, LaSB Procedures Manual, LSBSOP.030, Laboratory Services Branch (LaSB) Guidelines for the Determination and Documentation of Uncertainty of Chemical Measurements
- Standard Methods, APHA-AWWA-WEF, Standard Methods for the Examination of Water and Wastewater (21st Edition)
- Standard Methods for the Examination of Water and Wastewater (21st Edition): Volatile Organic Compounds – Purge & trap Capillary-Column Gas Chromatographic/Mass Spectrometric Method
- Standard Methods, Standard Methods News, Revised sections in Part 3000. 3500-Cr Chromium (2009), November 11, 2010 (www.standardmethods.org)
- United States Environmental Protection Agency (USEPA) Federal Register Part III. March 12, 2007. 40 CFR Part 122, 136, et al. Guidelines Establishing Test Procedures for the Analysis of Pollutants Under the Clean Water Act; National Primary Drinking Water Regulations; and National Secondary Drinking Water Regulations; Analysis and Sampling Procedures; Final Rule (page 11218, 11236, 11239)
- United States Environmental Protection Agency (USEPA) SW846. Test Methods for Evaluating Solid Waste, Physical/Chemical Methods
- United States Environmental Protection Agency (USEPA) Contract Laboratory Program, Office of Superfund Remediation and Technology Innovation. Multi-Media, Multi-concentration Organics Analysis, OLM04.3
- United States Environmental Protection Agency (USEPA) National Exposure Research Laboratory. Method 524.2: Measurement of Purgeable Organic Compounds in Water by Capillary Column Gas Chromatography/Mass Spectrometry
- United States Geological Survey (USGS), United States Department of the Interior
- United States Geological Survey (USGS) – National Water Quality Laboratory (NWQL)